

Real-time Detection of Pace Pulses in a Single Lead ECG

Irena Jekova¹, Serafim Tabakov², Ivo Iliev², Valentin Tsibulko², Krasimira Kostikova³

¹Institute of Biophysics and Biomedical Engineering, Bulgarian Academy of Sciences, Sofia, Bulgaria

²Technical University, Sofia, Bulgaria

³Medical University, Sofia, Bulgaria

Abstract

This paper presents an algorithm for real-time pace pulses detection in a single-lead ECG, based on assessment of the cumulative slope of the ECG signal calculated in a narrow time window. The algorithm is trained and tested on an artificial database containing 780 ECGs in lead II (390 for training, 390 for testing) that represent different arrhythmias, combined with artificially superimposed pace pulses, which cover the wide ranges of rising edge (<10 μ s to 100 μ s) and total pulse durations (100 μ s to 2 ms) and correspond to various pacemaker modes. The achieved accuracy is $Se=99.3\%$, $PPV=99.0\%$ for ECGs sampled at $Fs=32kHz$, and $Se=97.1\%$, $PPV=96.8\%$ for $Fs=16kHz$.

1. Introduction

Correct detection of pacemaker pulses in the electrocardiogram (ECG) is crucial for proper evaluation of its functionality and its effect on the cardiac rhythm, as well as for elimination of the pace pulses influence on the automatic ECG analysis.

The pace pulse is bipolar, with fast rising edge about 10 μ s; amplitude on the patient skin surface varying between few hundred μ V to several hundred mV; and width of pace artifacts between 100 μ s and 2ms [1]. Depending on the number of active leads the pacemakers are single chamber, dual chamber and bi-ventricular, while according to their programming the devices could be with fixed-rate, on demand and rate-responsive [2]. This inevitably leads to a number of challenges for the design of algorithms for pace pulses detection.

There are different medical standards with variable requirements regarding the height and width of the pace pulse that has to be captured and indicated on the screen of the device. According to ANSI/AAMI EC11 [3] the features of the pacemaker pulses that should be obligatory detected are as follows: (i) duration from 100 μ s to 2 ms; (ii) amplitude from 2 mV to 250 mV; (iii) frequency up to 100 pulses per minute; (iv) rising edge duration less than

100 ms. The IEC60601-2-27 standard [4] states duration from 0.5 ms to 2.0 ms and amplitude from 2 mV to 700 mV. Modern pacemakers could generate smaller pace pulse amplitudes that could fall below the requirements set in the standards and lead to complications in the algorithms for pace pulses detection [2].

Publicly presented pace pulse detection methods rely on hardware decisions [5] and software analysis after specific filtration [6,7] and multi-lead ECG processing [1]. The digital ECG analysis should be applied on high-resolution ECG [8,9] that preserves the frequency content of the pacing pulses.

This paper presents an algorithm for real-time pace pulses detection in a single ECG lead.

2. Database

The ECG signals used for training and testing are taken from an artificial database containing ECG recordings in lead II that represent different arrhythmias generated by HKP (Heidelberger Praxisklinik) simulator, combined with artificially superimposed pace pulses that cover the wide ranges of rising edge (<10 μ s to 100 μ s) and total pulse durations (100 μ s to 2 ms) and correspond to various pacemaker modes [10]. The database comprises 780 'pure' ECGs with pace pulses with duration of 10 s and annotated positions of the pace pulses. The signals are recorded with 9.81 μ V/LSB amplitude resolution at 128 kHz sampling rate, which preserves the steep raising and trailing edges of the pace pulses. The algorithm was trained on a set of 390 randomly selected recordings and tested on the remaining 390 ECGs.

3. Method

The algorithm for pace pulses detection is based on assessment of the cumulative slope of the ECG signal (*Slope*) calculated in a narrow time window (N samples), according to the equation:

$$Slope(j) = (2 * N * ECG(j) - \sum_{i=j-N}^{j+N} ECG(i))^2$$

Block-diagram of the algorithm is presented in Fig. 1.

$Slope(j)$ is compared to a falling threshold $SlopeTHR$ with predefined initial ($SlopeTHR_{Init}$) and minimal ($SlopeTHR_{Min}$) values. When $Slope(j)$ exceeds $SlopeTHR$ a pace pulse is detected, $SlopeTHR$ is updated and the detection of pace pulses is disabled for the next 10 ms in order to prevent consequent detections of the same pace pulse. After the detection procedure is unblocked, $SlopeTHR$ is slowly decreased with coefficient K until new pace pulse is detected or $SlopeTHR_{Min}$ is reached.




Figure 1. Block-diagram of the pace pulses detection algorithm.

We used the ECG recordings from the training dataset to assess the influence of the adjustable parameters N , $SlopeThr_{Init}$, $SlopeThr_{Min}$, K and the sampling frequency (F_s) on the reliability of pace pulses detection, and to select values, which could provide optimal detection accuracy.

3.1. Adjustment of N and $SlopeThr$

Two auxiliary measures, representing the maximal Slope value within 10 ms around the pace pulses annotations (MaxInPMann) and out of these intervals (MaxOutPMann), were calculated for ten different values of the time window N ($N = 0.1\text{ms}, 0.25\text{ms}, 0.5\text{ms}, 0.75\text{ms}, 1\text{ms}, 1.5\text{ms}, 2\text{ms}, 3\text{ms}, 4\text{ms}, 5\text{ms}$), as follows:

- $MaxInPMann(i) = \max(Slope(ANN(i) \pm 10\text{ms}))$
- $MaxOutPMann(i) = \max(Slope(ANN(i-1) + 10\text{ms} \text{ to } ANN(i) - 10\text{ms}))$

where i is the consecutive number of the pace pulse. Based on $MaxInPMann$, $MaxOutPMann$ values for each couple (N , $SlopeThr$) two statistical indices are defined:

- Sensitivity for detection of pace pulses:
 $Se = 100 * \text{count}(MaxInPMann \geq SlopeThr) / \text{count}(ANN)$

- Specificity for detection of pace pulses:

$$Sp = 100 * \text{count}(MaxOutPMann < SlopeThr) / \text{count}(ANN)$$

Receiver-operating characteristics (ROC) curves are used for selection of the optimal couple (N , $SlopeThr$).

3.2. Investigation of the influence of F_s

To assess the influence of F_s on the pace detection accuracy we calculated Se , Sp for the defined optimal time window N , different values of $SlopeThr$ and $F_s = 4, 8, 16, 32, 64, 128$ kHz. The maximal detection accuracy for each F_s was considered.

3.3. Selection of $SlopeThr_{Init}$, $SlopeThr_{Min}$ and adjustment of K

The ranges for adjustment of $SlopeThr_{Init}$ and $SlopeThr_{Min}$ were selected by considering the accuracy results for different fixed thresholds. Six K values were tested – $K(\%) = 1, 1/2, 1/4, 1/8, 1/16, 1/32$, i.e. $K(\%) = 10, 5, 2.5, 1.25, 0.625, 0.3125$.

The accuracy for the different combinations between $SlopeThr_{Init}$, $SlopeThr_{Min}$ and K was estimated via:

- Sensitivity: $Se = 100 * TP / (TP + FN)$
 - Positive predictive value: $PPV = 100 * TP / (TP + FP)$,
- where TP (true positive) is the number of the correctly detected pace pulses; FN (false negative) is the number of not detected pace pulses; FP (false positive) is the number of the erroneously detected pace pulses.

4. Results and discussion

The results achieved during the adjustment of the time window N and the optimal fixed value of the threshold $SlopeThr$ for selected $F_s=32$ kHz are presented via the ROC curves in Fig. 2. The ROC curves analysis highlights the following settings:

- $N=1\text{ms}$, $SlopeThr=1.28\text{V}$; $Sp=100\%$; $Se=99.28\%$;
- $N=1.5\text{ms}$, $SlopeThr=3.84\text{V}$; $Sp=100\%$; $Se=98.67\%$;
- $N=1.5\text{ms}$, $SlopeThr=2.35\text{V}$; $Sp=99.97\%$; $Se=99.52\%$;
- $N=1.5\text{ms}$, $SlopeThr=2.16\text{V}$; $Sp=97.04\%$; $Se=100\%$;

Based on the ROC curves in Fig. 2 a time window $N=1.5\text{ms}$ (providing balanced Se , Sp) is selected and the influence of F_s on the detection accuracy is investigated. The results are illustrated in Fig. 3. The ascending trend of the mean(Se, Sp) exceeds 99.5% for all $F_s \geq 32$ kHz. Mean(Se, Sp) falls down to 98 % for $F_s = 16$ kHz; 95.5 % for $F_s = 8$ kHz and 76.5 % for $F_s = 4$ kHz. Obviously, the choice of sampling frequency should be a compromise between detection accuracy on one side and the necessary processing resources and memory space for real-time analysis on the other.




Figure 2. ROC curves for different windows N and $SlopeThr$.




Figure 3. Influence of F_s on the detection accuracy.

Table 1. Accuracy of pace pulses detection for different F_s and $SlopeThr$

F_s (kHz)	$SlopeThr$ (V)	Sp (%)	Se (%)
32 kHz	3.82	100	98.67
	2.35	99.52	99.97
	2.16	97.04	100
16 kHz	1.32	100	96.26
	1.27	98.82	96.53
	1.18	96.41	96.92
	1.08	94.03	97.44
	0.98	86.88	97.98
	0.78	80.81	99.43
	0.59	79.34	100

Adjustment of initial threshold ($SlopeThrInit$), minimal threshold ($SlopeThrMin$) and coefficient K for $N=1.5$ ms, $F_s=32$ kHz and $F_s=16$ kHz is based on the accuracy results obtained for stationary threshold (Table 1).

Adjustments for $F_s=32$ kHz: We select $SlopeThrMin = 2.16$ V, for which $Se=100\%$. Lower values of $SlopeThrMin$ would only decrease Sp without any improvement of Se. Different K values are studied for $SlopeThrInit = 3.82$ V and $SlopeThrInit = 2.35$ V. The results are illustrated in Fig. 4a,b. The optimal combination of parameters for $F_s=32$ kHz is $SlopeThrInit = 3.82$ V, $SlopeThrMin = 2.16$ V, $K=0.3125\%$ with detection accuracy $Se=99.24\%$, $PPV=98.95\%$.

Adjustments for $F_s=16$ kHz: We select $SlopeThrMin = 1.08$ V, for which $Se=97.44\%$, $Sp=94.03\%$. Lower values of $SlopeThrMin$ would lead to insignificant increase of Se at the cost of considerable decrease in Sp. $SlopeThrInit$ is set to 1.32 V, for which $Sp=100\%$, and different K values are tested. The results are illustrated in Fig. 4c. Depending on the particular application and the preset requirements, for $F_s=16$ kHz one could select among (i) setting for maximal Se ($Se=97.09\%$, $PPV=96.75\%$) - $SlopeThrInit = 1.32$ V, $SlopeThrMin = 1.08$ V, $K=1.25\%$; (ii) setting for maximal PPV ($Se=96.25\%$, $PPV=97.86\%$) - $SlopeThrInit = 1.32$ V, $SlopeThrMin = 1.08$ V, $K=0\%$.




Fig. 4. Relation between Se (blue), PPV (red), mean(Se,PPV) (green) and the coefficient K for: (a) $SlopeThrInit=3.82$ V, $F_s=32$ kHz; (b) $SlopeThrInit=2.35$ V, $F_s=32$ kHz; (c) $SlopeThrInit = 1.32$ V, $F_s=16$ kHz.

The developed algorithm for pace pulses detection is tested in Matlab environment with the following settings:

- Setting 1: $F_s = 32$ kHz, $SlopeThrInit = 3.82$ V, $SlopeThrMin = 2.35$ V, $K = 0.3125\%$;
- Setting 2: $F_s = 16$ kHz, $SlopeThrInit = 1.32$ V, $SlopeThrMin = 1.08$ V, $K = 0\%$;
- Setting 3: $F_s = 16$ kHz, $SlopeThrInit = 1.32$ V, $SlopeThrMin = 1.08$ V, $K = 1.25\%$.

The test results are presented in Table 2.

Table 2. Accuracy of pace pulses detection over the test dataset, calculated for different settings.

Settings	Se (%)	PPV (%)
Setting 1	99.27	98.95
Setting 2	96.14	97.85
Setting 3	97.14	96.75

The operation of the developed algorithm with Setting 3 ($F_s = 16$ kHz, $SlopeThrInit = 1.32$ V, $SlopeThrMin = 1.08$ V, $K = 1.25\%$) is illustrated via the examples in figures 5 and 6. The false negative errors (decreasing Se) are due mainly to signals with low-amplitude pace pulses, while the false positive errors (decreasing Sp) result from steep and high-amplitude ectopic beats.




Fig. 5. Correct detection of low-amplitude pace pulses (red 'o' in 1st subplot mark the annotations, red '*' in 2nd subplot show the detections). The relatively high cumulative slope under the steep and high-amplitude ectopic beats does not lead to false pace pulses detection.




Fig. 6. Low-amplitude pace pulses not detected by the algorithm due to the low cumulative slope, which is comparable to the one for the ventricular complexes.

5. Conclusions

This study addresses the development and testing of an algorithm for detection of pace pulses based on quasi-real-time assessment of the cumulative ECG slope (delay of 1.5 ms). The designed method performs reliably and guarantees $Se=99.3\%$ and $PPV=99.0\%$ for single lead ECG sampled at $F_s=32$ kHz. Expectedly, when F_s is decreased to 16 kHz the detection accuracy drops down with 2 % ($Se=97.1\%$, $PPV=96.8\%$), generally due to false negative errors occurring for low-amplitude pace pulses.

Acknowledgements

This study is supported by the Bulgarian Scientific Research Fund, grant ДН17/19.

References

- [1] Kruse J., Redmond C. Detecting and distinguishing cardiac pacing artifacts. *Analog Dialogue* 2012;46:13-8.
- [2] Tsiulko V., Iliev I., Jekova I. A review on pacemakers: Device types, operating modes and pacing pulses. Problems related to the pacing pulses detection. *Int. J. Bioautomation* 2014;18(2):89-100.
- [3] American National Standard. Diagnostic electrocardiographic devices, ANSI/AAMI EC11, 2007.
- [4] American National Standard. Medical electrical equipment — Part 2-27: Particular requirements for the basic safety and essential performance of electrocardiographic monitoring equipment, ANSI/AAMI/IEC 60601-2-27, 2011
- [5] Calabria T. Hardware pace using slope detection. Texas Instruments Precision Designs: Verified Design, 2007.
- [6] Herleikson E. ECG pace pulse detection and processing. US5682902 A04-Nov-1997.
- [7] Polpetta A., Banelli P. Fully digital pacemaker detection in ECG signals using a non-linear filtering approach. Proc. 30th Annual International Conference of the IEEE Engineering in Medicine and Biology Society, EMBS'2008 2008:5406-10.
- [8] Petrutiu S., Sahakian A., Ricke A., Young B., Swiryn S. High resolution electrocardiography optimized for recording pulses from electronic pacemakers: Evaluation of a new pacemaker sensing system. *Computers in Cardiology* 2007;34:197-200.
- [9] Luo S., Johnston P., Hong W. Performance study of digital pacer spike detection as sampling rate changes. *Computing in Cardiology* 2008;35:349-52.
- [10] Jekova I., Tsiulko V., Iliev I. ECG Database Applicable for Development and Testing of Pace Detection Algorithms. *Int. J. Bioautomation* 2014;18(4):377-88.

Address for correspondence.

Irena Jekova
Bulgarian Academy of Sciences,
Institute of Biophysics and Biomedical Engineering,
Acad. G. Bonchev str., bl.105, 1113, Sofia, Bulgaria
irena@biomed.bas.bg