

Parameters	Measurement technique	Time resolution	Detection limit ^a	Accuracy
OH	LIF ^b	30 s	$4.0 \times 10^5 \text{ cm}^{-3}$	$\pm 14 \%$
HO ₂	LIF ^{b,c}	30 s	$1.0 \times 10^7 \text{ cm}^{-3}$	$\pm 17 \%$
RO ₂	LIF ^{b,c}	30 s	$0.5 \times 10^7 \text{ cm}^{-3}$	$\pm 23 \%$
k_{OH}	LP-LIF ^d	90 s	0.3 s^{-1}	$\pm 10 \% \pm 0.7 \text{ s}^{-1}$
Photolysis frequencies	Spectroradiometer	20 s	^e	$\pm 10 \%$
O ₃	UV photometry	60 s	0.5 ppbv	$\pm 5 \%$
NO	Chemiluminescence	60 s	60 pptv	$\pm 20 \%$
NO ₂	Chemiluminescence ^f	60 s	300 pptv	$\pm 20 \%$
HONO	LOPAP ^g	300 s	10 pptv	$\pm 20 \%$
CO, CH ₄ , CO ₂ , H ₂ O	CRDS	60 s	^h	ⁱ
SO ₂	Pulsed UV fluorescence	60 s	0.1 ppbv	$\pm 10 \%$
HCHO	Hantzsch fluorimetry	120 s	25 pptv	$\pm 5 \%$
Volatile organic compounds ^j	GC-FID/MS ^k	1 h	(20 ~ 300) pptv	$\pm (15 \sim 20) \%$
Oxygenated organic compounds	PTR-ToF-MS	10 s	(50–100) pptv	$\pm (10 \sim 15) \%$