



Effect of ions on the measurement of sulfuric acid in the CLOUD experiment at CERN

L. Rondo¹, A. Kürten¹, S. Ehrhart¹, S. Schobesberger², A. Franchin², H. Junninen², T. Petäjä², M. Sipilä², D. R. Worsnop^{2,3}, and J. Curtius¹

¹Institute for Atmospheric and Environmental Sciences, Goethe University Frankfurt am Main, 60438 Frankfurt am Main, Germany

²Department of Physics, University of Helsinki, P.O. Box 64, 00014 Helsinki, Finland

³Aerodyne Research, Inc., 45 Manning Road, Billerica, MA 01821, USA

Correspondence to: L. Rondo (l.rondo@iau.uni-frankfurt.de)

Received: 6 May 2014 – Published in Atmos. Meas. Tech. Discuss.: 3 July 2014

Revised: 26 September 2014 – Accepted: 30 September 2014 – Published: 19 November 2014

Abstract. Ternary aerosol nucleation experiments were conducted in the CLOUD chamber at CERN in order to investigate the influence of ions on new particle formation. Neutral and ion-induced nucleation experiments, i.e. without and with the presence of ions, respectively, were carried out under precisely controlled conditions. The sulfuric acid concentration was measured with a chemical ionisation mass spectrometer (CIMS) during the new particle formation experiments. The added ternary trace gases were ammonia (NH₃), dimethylamine (DMA, C₂H₇N) or oxidised products of pinanediol (PD, C₁₀H₁₈O₂). When pinanediol was introduced into the chamber, an increase in the mass spectrometric signal used to determine the sulfuric acid concentration (m/z 97, i.e. HSO₄⁻) was observed due to ions from the CLOUD chamber. The enhancement was only observed during ion-induced nucleation measurements by using either galactic cosmic rays (GCRs) or the proton synchrotron (PS) pion beam for the ion generation, respectively. The ion effect typically involved an increase in the apparent sulfuric acid concentration by a factor of ~ 2 to 3 and was qualitatively verified by the ion measurements with an atmospheric-pressure interface-time of flight (APi-TOF) mass spectrometer. By applying a high-voltage (HV) clearing field inside the CLOUD chamber, the ion effect on the CIMS measurement was completely eliminated since, under these conditions, small ions are swept from the chamber in about 1 s. In order to exclude the ion effect and to provide corrected sulfuric acid concentrations during the GCR and PS beam nucleation experiments, a parameterisation was derived that

utilises the trace gas concentrations and the UV light intensity as input parameters. Atmospheric sulfuric acid measurements with a CIMS showed an insignificant ion effect.

1 Introduction

Atmospheric aerosols have an important effect on clouds and climate, with secondary aerosol particles contributing significantly to the global concentration of cloud condensation nuclei (CCN) via new particle formation (Kulmala et al., 2004; Merikanto et al., 2010). From both field and laboratory measurements, it has been well established that sulfuric acid (H₂SO₄) plays a crucial role in atmospheric nucleation (Curtius, 2006; Riipinen et al., 2007; Kuang et al., 2008). However, binary nucleation of sulfuric acid and water vapour can be ruled out as a plausible mechanism to explain observed boundary layer nucleation rates (Kirkby et al., 2011). The presence of ternary substances like ammonia (Benson et al., 2009; Kirkby et al., 2011) or oxidised organic compounds (Zhang et al., 2004; Metzger et al., 2010; Riccobono et al., 2012, 2014; Schobesberger et al., 2013) can elevate the particle formation rates by orders of magnitude. Yet so far the physico-chemical processes taking place in atmospheric boundary layer particle formation remain unknown and questions still remain on the exact contribution of ternary substances and on the role of ions in atmospheric nucleation (Kulmala, 2003; Lovejoy et al., 2004). One goal of the CLOUD (Cosmics Leaving OUtdoor Droplets) experiment

at CERN is to investigate to what extent the presence of ions affects new particle formation at atmospherically relevant conditions (Kirkby et al., 2011; Schobesberger et al., 2013; Almeida et al., 2013). At CLOUD, neutral and ion-induced nucleation experiments are carried out in an aerosol chamber under precisely controlled conditions. Trace gases are added into the CLOUD chamber in order to investigate their influence on the nucleation. The source of ions in the chamber is either galactic cosmic rays (GCRs) or both GCR and CERN's proton synchrotron (PS) pion beam (Duplissy et al., 2010; Kirkby et al., 2011). Highly accurate and precise measurements of gaseous $[\text{H}_2\text{SO}_4]$ are required in order to interpret the data and for deriving parameterisations of the nucleation rate for certain conditions. Sulfuric acid concentration measurements are therefore performed with a chemical ionisation mass spectrometer (CIMS; Kürten et al., 2012).

In this paper, we describe the effect of ionising radiation on the sulfuric acid measurements performed by the CIMS instrument. During nucleation experiments with oxidation products of pinanediol, the mass spectrometric sulfuric acid signal (HSO_4^- ions) showed an increase due to the presence of ions in the chamber. The characteristics of the increased signal lead to the conclusion that HSO_4^- ions from the chamber were interfering with the product ions from the CIMS ion reaction zone. In order to verify this assumption, dedicated tests with the CIMS and the pion beam were performed. Surprisingly, the ion contribution was only detectable under the presence of oxidised organic compounds. Additionally, the atmospheric-pressure-interface time-of-flight mass spectrometer (APi-TOF; Junninen et al., 2010) measurements of ionic clusters confirmed that the enhanced HSO_4^- signal is coming directly from the chamber in the presence of oxidised organic compounds. In order to exclude the ion effect and to derive the correct sulfuric acid concentration for the ternary pinanediol nucleation runs, a parameterisation was developed.

2 Methods

2.1 CLOUD chamber measurements

Nucleation experiments were carried out using the CLOUD chamber at CERN (Kirkby et al., 2011). The electropolished stainless steel chamber of 26 m^3 volume provides an ultra-clean environment where selected trace gases and ultrapure humidified synthetic air can be continuously supplied. Well-defined amounts of H_2SO_4 are produced in situ from SO_2 and OH by means of internal ultraviolet (UV) illumination from a fibre-optic system producing OH from ozone and H_2O (Kupc et al., 2011). As a source of ionising radiation for ion-induced nucleation measurements, an adjustable pion beam from the CERN proton synchrotron (PS) is used in order to simulate and enhance the natural cosmic radiation. For the so-called GCR ion-induced nucleation experiments


(GCR runs), the pion beam is turned off and the chamber is only irradiated by the natural GCRs that enter the chamber, while for ion-free nucleation experiments (neutral runs), an internal electric clearing field, with a total potential difference of 60 kV, was used to sweep any ions out of the chamber within less than 1 s (Kirkby et al., 2011). The contents of the CLOUD chamber are continuously mixed by means of two fans (Voigtländer et al., 2012) and are analysed using a wide variety of instruments, each connected to a separate 0.75 m long sampling probe protruding into the chamber. The instrumentation includes several condensation particle counters (CPCs), mass spectrometers and electrical mobility analysers. Organic trace gases were monitored with a proton-transfer-reaction time-of-flight mass spectrometer (PTR-TOF-MS; Schnitzhofer et al., 2014), and aerosol particle concentrations were measured, starting at particle diameter sizes smaller than 2 nm (Wimmer et al., 2013).

During the CLOUD-04 campaign in June and July 2011, data of multi-component nucleation were obtained at the CLOUD chamber involving sulfuric acid and water vapour and adding either ammonia, dimethylamine or biogenic organic vapours (pinanediol, PD). PD is a surrogate first-generation product from the important biogenic vapour α -pinene, which is further oxidised by ozone and OH radicals to form more highly oxidised organic substances. Ammonia and dimethylamine concentrations were determined with an ion chromatograph (Praplan et al., 2012) while concentrations of organic compounds were measured with a PTR-TOF-MS (Schnitzhofer et al., 2014). The nucleation events were started by switching on the UV light, which initiates the OH formation from ozone photolysis and leads to the production of H_2SO_4 (Kupc et al., 2011). While pinanediol is present in the chamber, the produced OH radicals would also contribute to reactions with pinanediol, competing in this way with the sulfuric acid production (Schobesberger et al., 2013; Riccobono et al., 2014).

2.2 H_2SO_4 measuring technique

A negative-ion detection CIMS instrument was used to measure the concentration of gaseous H_2SO_4 with a detection limit of $\sim 1 \times 10^5$ molecule cm^{-3} and 30 s integration time. This system utilises a specific ion–molecule reaction to convert electrically neutral H_2SO_4 molecules into bisulfate HSO_4^- (m/z 97) ions, using NO_3^- as the ionising agent. This technique has been used previously in many other studies for gaseous or aerosol measurements (Viggiano and Arnold, 1983; Eisele and Tanner, 1993; Berresheim et al., 2000; Petäjä et al., 2009). The CIMS system (THS Instruments LLC, USA) contains a quadrupole mass filter and a Channeltron detector for the quantification of analyte ions (Huey, 2007). In addition to the basic components, a collision dissociation chamber (CDC) consisting of an octopole, is used to remove weakly bonded water and nitric acid molecules from the core ions via energetic collisions (Tanner et al., 1997).

The ion source that is used to generate the primary ions is a newly developed corona ion source (Kürten et al., 2011). For a known ion molecule residence time in the flow reactor and rate constant, the sulfuric acid concentration can be calculated by the ratio of the product ions (HSO_4^- ions) to educt ions (NO_3^- ions) count rates. In order to avoid any possible damage to the Channeltron detector due to the high count rate of the NO_3^- ions, we measure the isotope at m/z 64, $\text{N}^{18}\text{OO}_2^-$ (Kürten et al., 2012). The ion–molecule reactions between the reagent ions and the neutral compound to be detected and quantified occur in the flow reactor according to Reaction (1):



Subsequent to the formation of the bisulfate ion and nitric acid clusters within the CIMS flow reactor, the dissociation Reaction (R2) of the above bisulfate clusters takes place in the collision dissociation chamber, resulting in the detection of the HSO_4^- .



Electrostatic voltages applied to different parts of the ion source and the drift tube guide the primary ions into the sample flow, where they can interact with the H_2SO_4 molecules. At the end of the drift tube, primary and product ions are drawn into the vacuum chamber. Signals of the most prominent product and primary ions were recorded every few seconds, while a complete mass spectrum (up to $m/z \sim 250$) was recorded every few minutes, allowing for measurements in real time. For diffusion-controlled wall losses within the CIMS sampling line, the H_2SO_4 diffusion coefficient ($D = 0.078 \pm 0.0546 \text{ cm}^2 \text{ s}^{-1}$ for $T = 278 \text{ K}$ and $\text{RH} = 38 \%$; see Hanson and Eisele, 2000), the sample flow rate ($Q = 7.5 \pm 0.1 \text{ slpm}$) and the length of the sampling line ($L = 0.9 \text{ m}$) were taken into account. The transmission efficiency was calculated as 44 % for laminar flow in a straight tube using the empirical equations given in Baron and Willeke (2001). The measured H_2SO_4 concentration is derived by multiplying the $\text{HSO}_4^- / \text{NO}_3^-$ ratio by a calibration factor (Kürten et al., 2012). By performing a detailed calibration with a dedicated calibration source, uncertainties of reaction rate, reaction time and wall losses in the ion–molecule reactor do not influence the determination of the H_2SO_4 concentration. The calibration source is based on photolysing water vapour with a mercury lamp to generate a known and stable amount of OH radicals in front of the inlet. The produced OH radicals react with SO_2 from a gas bottle in the presence of O_2 and water vapour, resulting in known sulfuric acid concentrations. The CIMS instrument was calibrated before and just after the end of the CLOUD-04 campaign in order to assure high accuracy of the measurement.




Figure 1. (a) Measurement of the CIMS bisulfate ion signal (HSO_4^- , red trace) during a typical binary nucleation experiment, displayed along with the variation of the pion beam intensity (green) and the clearing field high voltage (blue). (b) Measurement of the CIMS bisulfate ion signal for the ternary system (involving pinenediol) under the same beam conditions as in (a) (magenta vertical lines show the ion effect under no-UV conditions). The CIMS signal is influenced by additional HSO_4^- ions from the chamber. This ion effect leads to an overestimation of the sulfuric acid concentration.

3 Results

3.1 Sulfuric acid measurements during binary nucleation experiments

The focus of the performed binary experiments ($\text{H}_2\text{SO}_4\text{--H}_2\text{O}$) was the measurement of the nucleation rate under neutral, GCR and charged pion beam conditions. With the initial experimental conditions established in the chamber with constant SO_2 and O_3 concentration, temperature (T), relative humidity (RH) and stable UV lamp operation, a stable sulfuric acid concentration is typically reached within about 15 min after opening the shutter of the UV light source. Depending on the H_2SO_4 concentration, aerosol nucleation occurs at a constant rate (Kirkby et al., 2011; Riccobono et al., 2014). After switching off the UV light, the OH production stops and the H_2SO_4 concentration decreases to a background level due to condensation of H_2SO_4 on the chamber walls and dilution with freshly supplied air (Voigtländer et al., 2012). An overview of a typical binary run is given in Fig. 1a, which displays the clearing field voltage, the UV light intensity, the pion beam





Figure 2. Enhancement of the HSO_4^- count rate in the presence of pinanediol (PD) during an experiment without the presence of UV light. The observed HSO_4^- enhancement in absence of UV light indicates the presence of a non-OH oxidant for SO_2 . The change in the PD concentration (factor of ~ 8) leads to an enhancement (factor of ~ 5) of the sulfuric acid background concentration (HSO_4^- count rate). For the period shown the SO_2 (3 ppbv) and O_3 (50 ppbv) mixing ratios remained stable; the clearing field high voltage was turned off during the first part of the experiment. The presence of ions generated by the pion beam (23:40 UTC and later) increased the CIMS signal by another factor of 5, confirming the ion effect in the presence of PD.

intensity and the HSO_4^- count rate used for deriving the sulfuric acid concentration. While the UV lights are turned off, there is a background of 5 to 7 counts per second for HSO_4^- , resulting in a background sulfuric acid concentration of 2.5×10^5 molecule cm^{-3} , which remains constant despite the changing clearing field voltage and pion beam intensity. Once the UV lights are switched on, the photooxidation of sulfur dioxide takes place in the presence of O_3 and H_2O , producing an equilibrium sulfuric acid concentration around 2×10^7 molecule cm^{-3} (HSO_4^- count rate ~ 300 s^{-1}). Similar to the situation under background conditions, the H_2SO_4 concentration is identical for neutral, GCR and charged pion beam conditions. After the UV lights are turned off, the H_2SO_4 steady state is disturbed and its concentration decreases due to loss onto the chamber walls and dilution. After some time (~ 1800 s), the H_2SO_4 equilibrates back to the initial background concentration that was measured before starting the nucleation run.

3.2 Sulfuric acid measurements during pinanediol nucleation experiments

In order to study the effect of oxidised organic compounds on the nucleation rates pinanediol (PD, $\text{C}_{10}\text{H}_{18}\text{O}_2$), vapour (Sigma Aldrich, 99 %) was introduced into the chamber (Schobesberger et al., 2013; Riccobono et al., 2014). The

concentration of the oxidation products (OxOrgs) is varied by changing the UV illumination and the concentration of pinanediol. The presence of PD (up to 500 ppbv) increased the HSO_4^- background intensity by as much as a factor of 40 in comparison to the chamber's neutral conditions without PD, i.e. during nucleation experiments using only H_2SO_4 and H_2O , or H_2SO_4 , H_2O and NH_3 (Fig. 1a and b). In particular, the enhancement of the HSO_4^- background signal is confirmed by the data shown in Fig. 2. During this experiment no UV light was present. However, the increase in the PD level resulted in a clear enhancement of the bisulfate signal. An additional increase in the HSO_4^- count rate is observed in the presence of the pion beam. The first increase in Fig. 2 suggests that another pathway exists for the production of H_2SO_4 besides the reaction between OH and SO_2 . It has been shown recently that stabilised Criegee intermediates produced from reactions between alkenes and ozone have the capacity to oxidise SO_2 (Mauldin III et al., 2012). However, since PD does not react with ozone due to its lack of a double bond, organic impurities (alkenes) (Schnitzhofer et al., 2014) fed into the chamber along with the PD could account not only for the production of Criegee intermediates but also for the production of OH via ozonolysis of these alkenes (Kroll et al., 2001) and the subsequent generation of sulfuric acid (Riccobono et al., 2014). The following simplified scheme summarises the possible reactions:



These organic impurities (mainly formaldehyde, acetaldehyde, acetone, formic acid and acetic acid), which have a rather high vapour pressure and are therefore not important for aerosol nucleation and growth, are reported to occur at rather low concentrations (below 1 ppbv in total; see Schnitzhofer et al., 2014). This dark production of sulfuric acid is not the focus of this study but will be studied in more detail in the future. Here we focus on the investigation of the ion effect on the CIMS sulfuric acid measurements in the presence of OxOrgs and ions (GCR and pion beam) for both background (UV off) and nucleation experimental runs (UV on).

An interesting feature in Fig. 1b is that, during background measurements in the presence of the pion beam (07:30 UTC and later), the HSO_4^- intensity measured by the mass spectrometer is increased by a factor of ~ 2 compared to neutral conditions (beam on or off but high voltage (HV) on). This observation indicates that there is an influence on the sulfuric acid measurement by additional HSO_4^- ions detected by the CIMS under these conditions (beam on and HV off). Thus, the CIMS is not only detecting the HSO_4^- ions being created in the CIMS flow reactor through reaction with NO_3^- ions but also detecting additional HSO_4^- ions coming directly from




Figure 3. Example of the effect of the pion beam on the HSO_4^- count rate during a nucleation experiment. During the second part of this experiment (16:00 UTC and later), the CIMS corona ion source was switched off and only ions from the chamber were detected.

the chamber. In addition, the same effect is also taking place in the presence of the UV irradiation, where the production of sulfuric acid is taking place through SO_2 oxidation by OH as well as stabilised Criegee intermediates (Riccobono et al., 2014). In both conditions (UV on and off), the observed enhancement of the detected HSO_4^- ions is about a factor of 2.

The formation of HSO_4^- ions within the chamber in the presence of GCRs or the pion beam is a result of ion–molecule reactions. The ions and free electrons created by the particle radiation rapidly interact with the gas molecules within the chamber and convert to complex positive and negative cluster ions and therefore induce radical chemical reactions (Kirkby, 2007). A chain of reactions is initiated by a rapid attachment of a free electron to an oxygen molecule O_2 which possesses an appreciable electron affinity, leading to O_2^- as the most important of the primary negative ions as shown in Reaction (R7). When an O_2^- ion collides with a sulfuric acid molecule, it charges the molecule or cluster negatively by removing one proton (Almeida et al., 2013).



By activating the clearing field (CF), the influence of the HSO_4^- ions from the chamber is directly confirmed (Fig. 1b). When the clearing field is switched on, the measured HSO_4^- signal drops by a factor of 2 within about 1 s, which corresponds to the time to remove ions due to the presence of the strong electric field (Kirkby et al., 2011). The H_2SO_4 lifetime of ~ 400 to 500 s is substantially longer; therefore, the observed short lifetime of the increased HSO_4^- signal confirms that ions are responsible for it and not neutral sulfuric acid.




Figure 4. APi-TOF negative ion mass spectrum for a binary H_2SO_4 – H_2O (blue) and a pinanediol (red) nucleation experiment with almost identical sulfuric acid concentration (6.6×10^6 molecule cm^{-3}). The HSO_4^- signal is increased by a factor of ~ 2 , which confirms that these ions are detected with a higher sensitivity if PD is present. Clustering with OxOrg molecules and subsequent fragmentation can explain the elevated HSO_4^- count rates.

In addition, a special experiment was performed by turning off the primary ion production in the CIMS corona ion source. The data in Fig. 3 unambiguously show that a substantial amount of HSO_4^- ions can be detected if the corona ion source is turned off while the pion beam is turned on.

Moreover, another vital confirmation of the ion effect is provided by the APi-TOF mass spectrometer measurements. The APi-TOF mass spectrometer measures the composition and concentrations of negative ions at mass-to-charge ratios ranging from about 50 to 3300 Th (Junninen et al., 2010; Schobesberger et al., 2013). A significant enhancement in the HSO_4^- signal measured by the APi-TOF is shown in Fig. 4 for the presence of oxidised organics. In this figure, selected ion and cluster ion signals are displayed both for a binary (H_2SO_4 – H_2O) and a ternary pinanediol nucleation experiment with very similar conditions. The concentration of H_2SO_4 is about 6.6×10^7 molecule cm^{-3} for both conditions, whereas the pion beam intensity is about 10 % higher for the PD experimental run. For the H_2SO_4 – H_2O nucleation experiment (binary run, shown in blue), the negative ion spectrum is clearly dominated by sulfuric acid and other inorganic compounds, which include water molecules associated with sulfuric acid or bisulfate. The strongest sulfate containing peak is from the HSO_4^- ion (m/z 97), which is seen to cluster with one, two or three H_2SO_4 molecules. The other main ions are HSO_4^- clustering with HNO_3 and other inorganic molecules. On the other hand, for the PD case (pinanediol run, shown in red), the most striking feature is that the HSO_4^- signal is raised by a factor of 2.2 even though the measured neutral sulfuric acid concentration is almost identical. Here,

the most abundant ions were complex cluster ions containing oxidised $C_{10}H_xO_y$ (where $x = 12, 14, 16$ and $y = 2-12$) organics and an HSO_4^- ion as well as 0–2 additional H_2SO_4 molecules.

4 Discussion

The APi-TOF mass spectrum clearly shows that HSO_4^- can cluster efficiently with oxidised organic products if PD is present in the chamber. In the H_2SO_4 – H_2O nucleation experiments these bisulfate–OxOrg clusters are absent and the mass spectrum is dominated by HSO_4^- . These observations suggest that the presence of OxOrg can enhance the transmission efficiency of the HSO_4^- core ions through the sampling line of the APi-TOF and the CIMS. In particular, the sampling line transmission efficiency due to diffusional losses (Baron and Willeke, 2001) varies from 59 to 69 % for OxOrg · HSO_4^- and OxOrg · $(H_2SO_4)_3 \cdot HSO_4^-$, respectively, while the transmission efficiencies for the CIMS sampling line are 64 % and 72 % for the same ions. The increase in the mass of the $HSO_4^- \cdot OxOrg$ ion complex leads to a reduction in the ion mobility, thereby enhancing the probability that the $HSO_4^- \cdot OxOrg$ ions can eventually be detected. Fragmentation (i.e. evaporation of the OxOrg ligand) in the mass spectrometer leads to a rather strong increase in the HSO_4^- signal. Since the CIMS uses a collision dissociation chamber (CDC) to detect only the core product ions and the mass range is limited to $\sim m/z$ 250, no mixed $HSO_4^- \cdot OxOrg$ clusters are visible in the mass spectrum (see Fig. 5). The APi-TOF, however, appears to fragment the ions less strongly than the CIMS does and therefore the mixed $HSO_4^- \cdot OxOrg$ clusters are clearly detected. A closer look at the sampling conditions of the CIMS reveals that, for conditions without organics, a major part of the bisulfate ions from the chamber is probably being lost before reaching the ion drift tube due to diffusion or electrostatic forces. Although there is no direct experimental evidence that ions are lost at one specific location in the CIMS sampling line and ion drift region, we can presume several possible locations for such effects to occur (see Fig. 6): (1) the entrance of the sampling probe, (2) the inner surface of the sampling line, (3) the section where the sampling line terminates and transitions into the CIMS drift tube and (4) the entrance to the vacuum chamber of the mass spectrometer (pinhole plate). Loss due to effects 3 and 4 are probably dominant for ions because electric fields separate the different regions of the CIMS inlet part (Kürten et al., 2011). Negatively charged ions in the sample gas experience a repulsing field when they enter the ion drift tube, and ions with high mobility can very likely be precipitated efficiently at the end of the sampling line. This loss mechanism is likely less strongly pronounced when OxOrgs are bound to HSO_4^- and make the ion cluster less mobile. However, during binary and ternary (NH_3 , DMA) experiments, data from the APi-TOF show that HSO_4^- does not bind with DMA or




Figure 5. CIMS negative mass spectrum for a binary H_2SO_4 – H_2O (blue) and a pinanediol (red) experimental run in the presence of ions. For the pinanediol experimental run, the HSO_4^- signal is enhanced by a factor of 2.25 in comparison to the equivalent neutral run (green).

NH_3 molecules (Schobesberger et al., 2013). Therefore, in the presence of these gases there is no enhancement in the HSO_4^- detection efficiency with the CIMS.

5 Parameterisation of H_2SO_4 concentration during pinanediol experiments

From the discussion above it is clear that ions had an effect on the measured sulfuric acid concentrations during ternary ion-induced nucleation experiments involving pinanediol. Only during neutral runs, while the high voltage of the clearing field electrodes is enabled, can the detected HSO_4^- ion signal be directly used to determine the sulfuric acid concentration in the chamber. Any ion effect can be completely excluded as an additional source of HSO_4^- ions, since under these conditions small ions are swept from the chamber within less than 1 s. However, in order to derive the correct sulfuric acid concentration during the ion-induced nucleation experiments, one needs to be aware of ion effects and correct for them. Because not every ion-induced nucleation run was preceded by a neutral run with identical conditions, a parameterisation was derived, which can be applied to all ion-induced nucleation runs with pinanediol in order to derive a corrected sulfuric acid concentration. The parameterisation is only a simplification of the real sulfuric acid production process and it is not based on a physico-chemical modelled process. The chemistry behind the modelled processes is complicated, and not all affecting factors can be fully taken into account. For example, reactions in which SO_2 is oxidised into sulfuric acid in the presence of UV radiation are known but the stabilised Criegee intermediates and the species involved, like




Figure 6. Schematic of the CLOUD chamber and the connected CIMS mass spectrometer. In a typical ion-induced binary (i.e. $\text{H}_2\text{SO}_4 + \text{H}_2\text{O}$) experiment (HV is switched off and the pion beam is on), not only H_2SO_4 neutral molecules but also HSO_4^- ions are produced, which are usually attached to water molecules (yellow), while for pinanediol runs the HSO_4^- ions are very likely bound to bigger OxOrg clusters (magenta). During binary experiments, the produced bisulfate ions are presumably efficiently lost before reaching the mass spectrometer. Losses can occur due to diffusion and/or electrostatic forces: (1) at the entrance of the sampling probe, (2) within the sampling line, (3) during the transfer into the CIMS ion drift region and (4) before entering the mass spectrometer.

OH radicals, also interact with several other compounds in the chamber, thus making it difficult to estimate how much sulfuric acid is actually formed when OH is not directly measured. In addition, DMA chemistry should also be considered since it reacts not only with O_3 , producing OH radicals (Tuaizon et al., 1994), but also consumes OH radicals (Atkinson et al., 1978). The parameterisation provides an estimate of the net effect of these sinks and sources acting in the CLOUD chamber. The results of the parameterisation indicate that the derived $[\text{H}_2\text{SO}_4]$ is able to represent the “true” sulfuric acid concentration adequately. For the parameterisation of the sulfuric acid, a nonlinear regression analysis was used in this study in order to find a predictive model for the sulfuric acid concentration. Figure 7 shows that the predictive ability of the derived $[\text{H}_2\text{SO}_4]$ in an ion-free environment is good (red points), whereas for the ion-induced pinanediol case (green points) the observed deviation indicates the ion effect on the CIMS measurements. The developed parameterisation includes a combination of measured O_3 , SO_2 , relative humidity (RH), UV intensity, dimethylamine (DMA) and pinanediol (PD) mixing ratios. In more detail, the parameterisation relies on the following formula:

$$[\text{H}_2\text{SO}_4]_{\text{calc},i} = \text{Bg}_i + \left(a_0 + a_1 \cdot \text{UV}_i + a_2 \cdot \text{UV}_i^2 + a_3 \cdot \text{UV}_i^3 \right) \cdot \left(k \cdot \text{RH}_i^b \cdot \text{SO}_{2,i}^c \cdot \text{O}_{3,i}^d + g \cdot \text{PD}_i^j + m \cdot \text{DMA}_i^l \right).$$

Here, the index i represents an averaged value for one experimental run denoting each of the parameters, UV represents the aperture of the UV light system in percent and RH represents the relative humidity in percent (temperature was constant at 278 K). As for the gases used, SO_2 is the sulfur dioxide mixing ratio in ppbv, O_3 the ozone mixing ratio in ppbv, PD the pinanediol mixing ratio in pptv, DMA the dimethylamine mixing ratio in pptv and Bg the background concentration of sulfuric acid just before an experimental run. Fitting the equation to the experimental data with the software IGOR yields the coefficients shown in Table 1 (coefficient values $\pm 95\%$ confidence interval). From the comparison between parameterised and experimentally determined H_2SO_4 concentrations for the neutral runs, it is shown that the data are in good agreement ($R^2 = 0.963$) (Fig. 7), which indicates that the parameterisation with a standard error of $\pm 6.3 \times 10^5 \text{ cm}^{-3}$ can also be applied to the GCR and charged beam experimental runs when no information on the $[\text{H}_2\text{SO}_4]$ under neutral conditions is available. In order to confirm the good agreement between the




Figure 7. Calculated $[\text{H}_2\text{SO}_4]$ vs. measured $[\text{H}_2\text{SO}_4]$ for binary and ternary organic nucleation runs. The presence of ions during the pinanediol runs leads to an excess of measured H_2SO_4 . The deviation between the parameterised and the measured concentrations for the charged runs indicates the magnitude of the ion effect on the CIMS measurement.

Table 1. Coefficients (\pm SD) of parameters used in the H_2SO_4 parameterisation formula.

Parameter	Value
k	$2.3 \times 10^{-4} \pm 1.4 \times 10^{-6}$
a_0	$51\,593 \pm 1$
a_1	$12\,953 \pm 1$
a_2	-209.99 ± 0.001
a_3	$1.0934 \pm 1 \times 10^{-5}$
b	$2.0934 \pm 1 \times 10^{-5}$
c	0.82661 ± 0.127
d	0.56574 ± 0.0869
g	$-1.9897 \times 10^{-5} \pm 0.00025$
m	-0.747 ± 2.58
j	1.2766 ± 1.25
l	-0.087045 ± 1

parameterised and experimental data, in Fig. 8 the experimental neutral runs are displayed along with their equivalent GCR and pion beam experimental runs that followed in sequence. The parameterised sulfuric acid concentration was of advantage not only for the current study but also for further analysis of nucleation studies from the CLOUD experiments (Riccobono et al., 2014).

6 Atmospheric implications

A question related to the discussed observations is to what degree atmospheric measurements could be affected by this




Figure 8. For some GCR (galactic cosmic ray) and pion beam runs a neutral run preceded where the conditions of the chamber were identical. This figure compares the parameterised $[\text{H}_2\text{SO}_4]$ for the GCR and beam runs to the value from the neutral stage. The good agreement indicates that the parameterisation can also be applied to the GCR and beam runs when no information on the $[\text{H}_2\text{SO}_4]$ under neutral conditions is available.

phenomenon. Because the CLOUD-04 experiment simulated atmospherically relevant oxidation products to a great extent (Riccobono et al., 2014), there was a need for further investigation of the ambient bisulfate ion enhancement effect on the CIMS measurements. As a consequence of our observations, we developed an ion precipitator, which was installed inside of a new ambient sampling line (total length 1 m and flow rate 1800 L min^{-1}) and was located just before the entrance to the ion source. The ion precipitator uses a high voltage in the range of 2 kV and an applied grounding located opposite to the high voltage, resulting in an electrostatic field which effectively removes ions from the ambient sample flow. This adjustment in the CIMS sampling line was developed in order to investigate the effect of ions on atmospheric measurements of gaseous sulfuric acid and for future CLOUD experiments to avoid the additional signal of HSO_4^- ions coming from the chamber.

During August and September 2011, the CIMS instrument participated in the PARADE campaign, which took place at the Kleiner Feldberg Taunus Observatory near Frankfurt, Germany. Figure 9 shows the measurement of sulfuric acid during daytime. In the lower part of the graph the sulfuric acid concentration is displayed along with the ion precipitator setting. The ion precipitator is set to be active every 15 min in order to investigate whether an effect from ambient ions can be detected. The performed measurements do not indicate a significant influence of the ion effect, and we conclude that the contribution of the atmospheric HSO_4^- ions is very small during the PARADE measurements. Turning off




Figure 9. Daytime atmospheric $[H_2SO_4]$ measurements along with the setting of the ion precipitator. In the upper graph, the count rates of the primary ions NO_3^- and the HSO_4^- product ions are displayed while the ion precipitator is enabled every 15 min. While the corona ion source is on, the signal of both NO_3^- and HSO_4^- is not noticeably influenced by the ion precipitator. When the CIMS corona ion source is off, the maximum contribution of the ambient HSO_4^- ions can be evaluated to be several 10^4 molecule cm^{-3} .

the corona ion source allows for estimation of the maximum contribution due to ambient ions (lower panel of Fig. 9). If the ion precipitator is enabled, all bisulfate ions are effectively eliminated (0 counts s^{-1}). In contrast, for the case of having both the corona source and the ion precipitator switched off, the slight increase in the bisulfate ion signal (~ 4 to 5 counts s^{-1}) is noticeably observed relative to the previous case. Using this measured HSO_4^- ion count rate together with the typical primary ion count rate, one can translate the effect of the ambient ions to be equivalent to several 10^4 molecule cm^{-3} during normal operation. This is typically smaller than the CIMS noise level.

Nevertheless, earlier studies have shown, for example, that significant ion production can occur in the vicinity of high-voltage power lines (Eisele, 1989a, b; Matthews et al., 2010). Therefore, in other environments with a strong local source of ions, it is not clear whether CIMS measurements could potentially be affected more strongly. Still, during the entire PARADE campaign, the GCR ion effect was negligible for the CIMS H_2SO_4 concentration measurements. Hence, since the exact nature of the bisulfate transmission enhancing ligand that is leading to the enhanced ion detection has not yet been identified, for other environments, sampling conditions, ion source geometries and the applied electric fields, it cannot be ruled out that sulfuric acid atmospheric measurements or chamber experiments are affected to some extent.

7 Conclusions

In this study the effect of ions on CIMS sulfuric acid concentration measurements in the presence of the organic vapour pinenediol (PD) and its oxidised products has been presented. We measured the sulfuric acid concentration during the CLOUD-04 campaign, in experiments studying nucleation from $H_2SO_4-H_2O$ (binary) and $H_2SO_4-H_2O-NH_3$, $H_2SO_4-H_2O-DMA$, $H_2SO_4-H_2O-DMA-PD$ or $H_2SO_4-H_2O-PD$ (ternary) systems of vapours within the CLOUD chamber, where all parameters of the experiment were well controlled. In the presence of pinenediol organic vapour and its oxidised products, and while ions are introduced into the chamber, there is an amount of HSO_4^- produced within the chamber which is effectively detected, leading to an overestimation of the H_2SO_4 monomer concentration by the CIMS measurement. We performed tests of the CIMS instrument in order to verify that we count more HSO_4^- ions in the presence of chamber ions. Moreover, data from the API-TOF confirmed the enhanced detection efficiency of sulfuric acid ions, indicating a ligand that decreases the mobility of the bisulfate ion complex. As the HSO_4^- ions affect the CIMS H_2SO_4 measurement, we developed a parameterisation of the CIMS sulfuric acid measurements for the neutral case, when all ions are removed from the chamber within < 1 s and therefore no HSO_4^- is present in the chamber. Later on, the parameterisation was applied to the CIMS sulfuric acid measurements during GCR and charged pion beam experimental runs. In this way, we corrected the CIMS sulfuric acid measurements for the ion effect. We were also interested in the atmospheric implications of this observation, so for this reason we performed atmospheric sulfuric acid measurements with the CIMS mass spectrometer, which showed that the ion effect was negligible for ground-based ambient measurements. Nevertheless, for future chamber experiments, an ion precipitator should in general be applied in front of the CIMS for regular operation. Furthermore, more dedicated studies should be carried out to investigate the ion effect of complex bisulfate ions on the sulfuric acid CIMS measurements in the presence of oxidised organics, focusing on the transmission efficiency of the bisulfate ions to the mass spectrometer.

Acknowledgements. We thank CERN for supporting CLOUD with important technical and financial resources and provision of a particle beam from the proton synchrotron. This research was funded by the EC 7th Framework Programme (Marie Curie Initial Training Network “CLOUD-ITN”, grant no. 215072), the German Federal Ministry of Education and Research (project no. 01LK0902A), the ERC Advanced Grant “ATMNUCLE” (no. 227463), and the Academy of Finland via the Centre of Excellence programme (project no. 1118615, 272041) and grant no. 1133872, 272851.

Edited by: J. Abbatt

References

- Almeida, J., Schobesberger, S., Kürten, A., Ortega, I. K., Kupiainen-Määttä, O., Praplan, A. P., Adamov, A., Amorim, A., Bianchi, F., Breitenlechner, M., David, A., Dommen, J., Donahue, N. M., Downard, A., Dunne, E., Duplissy, J., Ehrhart, S., Flagan, R. C., Franchin, A., Guida, R., Hakala, J., Hansel, A., Heinritzi, M., Henschel, H., Jokinen, T., Junninen, H., Kajos, M., Kangasluoma, J., Keskinen, H., Kupc, A., Kurtén, T., Kvashin, A. N., Laaksonen, A., Lehtipalo, K., Leiminger, M., Leppä, J., Loukonen, V., Makhmutov, V., Mathot, S., McGrath, M. J., Nieminen, T., Olenius, T., Onnela, A., Petäjä, T., Riccobono, F., Riipinen, I., Rissanen, M., Rondo, L., Ruuskanen, T., Santos, F. D., Sarnela, N., Schallhart, S., Schnitzhofer, R., Seinfeld, J. H., Simon, M., Sipilä, M., Stozhkov, Y., Stratmann, F., Tomé, A., Tröstl, J., Tsigkogeorgas, G., Vaattovaara, P., Viisanen, Y., Virtanen, A., Vrtala, A., Wagner, P. E., Weingartner, E., Wex, H., Williamson, C., Wimmer, D., Ye, P., Yli-Juuti, T., Carslaw, K. S., Kulmala, M., Curtius, J., Baltensperger, U., Worsnop, D. R., Vehkamäki, H., and Kirkby, J.: Molecular understanding of sulphuric acid-amine particle nucleation in the atmosphere, *Nature*, 502, 359–363, doi:10.1038/nature12663, 2013.
- Atkinson, R., Perry, R. A., Pitts Jr., J. N.: Rate constants for the reactions of the hydroxyl radical with dimethylamine, trimethylamine, and ethylamine over the temperature range 298–426 K, *J. Chem. Phys.*, 68, 1850–1853, 1978.
- Baron, P. A. and Willeke, K.: *Aerosol Measurement: Principles, Techniques, and Applications*, John Wiley and Sons, 2nd Edn., p. 580, 2001.
- Benson, D. R., Erupe, M. E., and Lee, S.-H.: Laboratory-measured $\text{H}_2\text{SO}_4\text{-H}_2\text{O-NH}_3$ ternary homogeneous nucleation rates: Initial observations, *Geophys. Res. Lett.*, 36, L15818, doi:10.1029/2009GL038728, 2009.
- Berresheim, H., Elste, T., Plass-Dülmer, C., Eisele, F. L., and Tanner, D. J.: Chemical ionization mass spectrometer for long-term measurements of atmospheric OH and H_2SO_4 , *Int. J. Mass Spectrom.*, 202, 91–109, doi:10.1016/S1387-3806(00)00233-5, 2000.
- Curtius, J.: Nucleation of atmospheric aerosol particles, *Comptes Rendus Phys.*, 7, 1027–1045, doi:10.1016/j.crhy.2006.10.018, 2006.
- Duplissy, J., Enghoff, M. B., Aplin, K. L., Arnold, F., Aufmhoff, H., Avngaard, M., Baltensperger, U., Bondo, T., Bingham, R., Carslaw, K., Curtius, J., David, A., Fastrup, B., Gagné, S., Hahn, F., Harrison, R. G., Kellett, B., Kirkby, J., Kulmala, M., Laakso, L., Laaksonen, A., Lillestol, E., Lockwood, M., Mäkelä, J., Makhmutov, V., Marsh, N. D., Nieminen, T., Onnela, A., Pedersen, E., Pedersen, J. O. P., Polny, J., Reichl, U., Seinfeld, J. H., Sipilä, M., Stozhkov, Y., Stratmann, F., Svensmark, H., Svensmark, J., Veenhof, R., Verheggen, B., Viisanen, Y., Wagner, P. E., Wehrle, G., Weingartner, E., Wex, H., Wilhelmsson, M., and Winkler, P. M.: Results from the CERN pilot CLOUD experiment, *Atmos. Chem. Phys.*, 10, 1635–1647, doi:10.5194/acp-10-1635-2010, 2010.
- Eisele, F. L.: Natural and atmospheric negative ions in the troposphere, *J. Geophys. Res.-Atmos.*, 94, 2183–2196, 1989a.
- Eisele, F. L.: Natural and transmission line produced positive ions, *J. Geophys. Res.-Atmos.*, 94, 6309–6318, 1989b.
- Eisele, F. L. and Tanner, D. J.: Measurement of the gas phase concentration of H_2SO_4 and methane sulfonic acid and estimates of H_2SO_4 production and loss in the atmosphere, *J. Geophys. Res.-Atmos.*, 98, 9001–9010, doi:10.1029/93JD00031, 1993.
- Hanson, D. R. and Eisele, F. L.: Diffusion of H_2SO_4 in Humidified Nitrogen: Hydrated H_2SO_4 , *J. Phys. Chem. A*, 104, 1715–1719, 2000.
- Huey, L. G.: Measurement of trace atmospheric species by chemical ionization mass spectrometry: Speciation of reactive nitrogen and future directions, *Mass Spectrom. Rev.*, 26, 166–184, doi:10.1002/mas.20118, 2007.
- Junninen, H., Ehn, M., Petäjä, T., Luosujärvi, L., Kotiaho, T., Koskiainen, R., Rohner, U., Gonin, M., Fuhrer, K., Kulmala, M., and Worsnop, D. R.: A high-resolution mass spectrometer to measure atmospheric ion composition, *Atmos. Meas. Tech.*, 3, 1039–1053, doi:10.5194/amt-3-1039-2010, 2010.
- Kirkby, J.: Cosmic rays and climate, *Surv. Geophys.*, 28, 333–375, 2007.
- Kirkby, J., Curtius, J., Almeida, J., Dunne, E., Duplissy, J., Ehrhart, S., Franchin, A., Gagné, S., Ickes, L., Kürten, A., Kupc, A., Metzger, A., Riccobono, F., Rondo, L., Schobesberger, S., Tsigkogeorgas, G., Wimmer, D., Amorim, A., Bianchi, F., Breitenlechner, M., David, A., Dommen, J., Downard, A., Ehn, M., Flagan, R. C., Haider, S., Hansel, A., Hauser, D., Jud, W., Junninen, H., Kreissl, F., Kvashin, A., Laaksonen, A., Lehtipalo, K., Lima, J., Lovejoy, E. R., Makhmutov, V., Mathot, S., Mikkilä, J., Minginette, P., Mogo, S., Nieminen, T., Onnela, A., Pereira, P., Petäjä, T., Schnitzhofer, R., Seinfeld, J. H., Sipilä, M., Stozhkov, Y., Stratmann, F., Tomé, A., Vanhanen, J., Viisanen, Y., Vrtala, A., Wagner, P. E., Walther, H., Weingartner, E., Wex, H., Winkler, P. M., Carslaw, K. S., Worsnop, D. R., Baltensperger, U., and Kulmala, M.: Role of sulphuric acid, ammonia and galactic cosmic rays in atmospheric aerosol nucleation, *Nature*, 476, 429–433, doi:10.1038/nature10343, 2011.
- Kroll, J. H., Sahay, S. R., Anderson, J. G., Demerjian, K. L., and Donahue, N. M.: Mechanism of HOx formation in the gas-phase ozone-alkene reaction. 2. Prompt versus thermal dissociation of carbonyl oxides to form OH, *J. Phys. Chem. A*, 105, 4446–4457, 2001.
- Kuang, C., McMurry, P. H., McCormick, A. V., and Eisele, F. L.: Dependence of nucleation rates on sulfuric acid vapor concentration in diverse atmospheric locations, *J. Geophys. Res.-Atmos.*, 113, D10209, doi:10.1029/2007JD009253, 2008.
- Kulmala, M.: How Particles Nucleate and Grow, *Science*, 302, 1000–1001, doi:10.1126/science.1090848, 2003.
- Kulmala, M., Vehkamäki, H., Petäjä, T., Dal Maso, M., Lauri, A., Kerminen, V.-M., Birmili, W., and McMurry, P. H.: Formation and growth rates of ultrafine atmospheric particles: a review of observations, *J. Aerosol Sci.*, 35, 143–176, doi:10.1016/j.jaerosci.2003.10.003, 2004.
- Kupc, A., Amorim, A., Curtius, J., Danielczok, A., Duplissy, J., Ehrhart, S., Walther, H., Ickes, L., Kirkby, J., Kürten, A., Lima, J. M., Mathot, S., Minginette, P., Onnela, A., Rondo, L., and Wagner, P. E.: A fibre-optic UV system for H_2SO_4 production in aerosol chambers causing minimal thermal effects, *J. Aerosol Sci.*, 42, 532–543, doi:10.1016/j.jaerosci.2011.05.001, 2011.
- Kürten, A., Rondo, L., Ehrhart, S., and Curtius, J.: Performance of a corona ion source for measurement of sulfuric acid by chemical ionization mass spectrometry, *Atmos. Meas. Tech.*, 4, 437–443, doi:10.5194/amt-4-437-2011, 2011.

- Kürten, A., Rondo, L., Ehrhart, S., and Curtius, J.: Calibration of a Chemical Ionization Mass Spectrometer for the Measurement of Gaseous Sulfuric Acid, *J. Phys. Chem. A*, 116, 6375–6386, doi:10.1021/jp212123n, 2012.
- Lovejoy, E. R., Curtius, J., and Froyd, K. D.: Atmospheric ion-induced nucleation of sulfuric acid and water, *J. Geophys. Res.-Atmos.*, 109, D08204, doi:10.1029/2003JD004460, 2004.
- Matthews, J. C., Ward, J. P., Keitch, P. A., and Henshaw, D. L.: Corona ion induced atmospheric potential gradient perturbations near high voltage power lines, *Atmos. Environ.*, 44, 5093–5100, doi:10.1016/j.atmosenv.2010.09.007, 2010.
- Mauldin III, R. L., Berndt, T., Sipilä, M., Paasonen, P., Petäjä, T., Kim, S., Kurtén, T., Stratmann, F., Kerminen, V.-M., and Kulmala, M.: A new atmospherically relevant oxidant of sulphur dioxide, *Nature*, 488, 193–196, doi:10.1038/nature11278, 2012.
- Merikanto, J., Spracklen, D. V., Pringle, K. J., and Carslaw, K. S.: Effects of boundary layer particle formation on cloud droplet number and changes in cloud albedo from 1850 to 2000, *Atmos. Chem. Phys.*, 10, 695–705, doi:10.5194/acp-10-695-2010, 2010.
- Metzger, A., Verheggen, B., Dommen, J., Duplissy, J., Prevot, A. S. H., Weingartner, E., Riipinen, I., Kulmala, M., Spracklen, D. V., Carslaw, K. S., and Baltensperger, U.: Evidence for the role of organics in aerosol particle formation under atmospheric conditions, *Proc. Natl. Acad. Sci.*, 107, 6646–6651, doi:10.1073/pnas.0911330107, 2010.
- Petäjä, T., Mauldin, III, R. L., Kosciuch, E., McGrath, J., Nieminen, T., Paasonen, P., Boy, M., Adamov, A., Kotiaho, T., and Kulmala, M.: Sulfuric acid and OH concentrations in a boreal forest site, *Atmos. Chem. Phys.*, 9, 7435–7448, doi:10.5194/acp-9-7435-2009, 2009.
- Praplan, A. P., Bianchi, F., Dommen, J., and Baltensperger, U.: Dimethylamine and ammonia measurements with ion chromatography during the CLOUD4 campaign, *Atmos. Meas. Tech.*, 5, 2161–2167, doi:10.5194/amt-5-2161-2012, 2012.
- Riccobono, F., Rondo, L., Sipilä, M., Barmet, P., Curtius, J., Dommen, J., Ehn, M., Ehrhart, S., Kulmala, M., Kürten, A., Mikkilä, J., Paasonen, P., Petäjä, T., Weingartner, E., and Baltensperger, U.: Contribution of sulfuric acid and oxidized organic compounds to particle formation and growth, *Atmos. Chem. Phys.*, 12, 9427–9439, doi:10.5194/acp-12-9427-2012, 2012.
- Riccobono, F., Schobesberger, S., Scott, C. E., Dommen, J., Ortega, I. K., Rondo, L., Almeida, J., Amorim, A., Bianchi, F., Breitenlechner, M., David, A., Downard, A., Dunne, E. M., Duplissy, J., Ehrhart, S., Flagan, R. C., Franchin, A., Hansel, A., Junninen, H., Kajos, M., Keskinen, H., Kupc, A., Kürten, A., Kvashin, A. N., Laaksonen, A., Lehtipalo, K., Makhmutov, V., Mathot, S., Nieminen, T., Onnela, A., Petäjä, T., Praplan, A. P., Santos, F. D., Schallhart, S., Seinfeld, J. H., Sipilä, M., Spracklen, D. V., Stozhkov, Y., Stratmann, F., Tomé, A., Tsagkogeorgas, G., Vaattovaara, P., Viisanen, Y., Vrtala, A., Wagner, P. E., Weingartner, E., Wex, H., Wimmer, D., Carslaw, K. S., Curtius, J., Donahue, N. M., Kirkby, J., Kulmala, M., Worsnop, D. R., and Baltensperger, U.: Oxidation products of biogenic emissions contribute to nucleation of atmospheric particles, *Science*, 344, 717–721, doi:10.1126/science.1243527, 2014.
- Riipinen, I., Sihto, S.-L., Kulmala, M., Arnold, F., Dal Maso, M., Birmili, W., Saarnio, K., Teinilä, K., Kerminen, V.-M., Laaksonen, A., and Lehtinen, K. E. J.: Connections between atmospheric sulphuric acid and new particle formation during QUEST III-IV campaigns in Heidelberg and Hyytiälä, *Atmos. Chem. Phys.*, 7, 1899–1914, doi:10.5194/acp-7-1899-2007, 2007.
- Schnitzhofer, R., Metzger, A., Breitenlechner, M., Jud, W., Heinrich, M., De Menezes, L.-P., Duplissy, J., Guida, R., Haider, S., Kirkby, J., Mathot, S., Minginette, P., Onnela, A., Walther, H., Wasem, A., Hansel, A., and the CLOUD Team: Characterisation of organic contaminants in the CLOUD chamber at CERN, *Atmos. Meas. Tech.*, 7, 2159–2168, doi:10.5194/amt-7-2159-2014, 2014.
- Schobesberger, S., Junninen, H., Bianchi, F., Lönn, G., Ehn, M., Lehtipalo, K., Dommen, J., Ehrhart, S., Ortega, I. K., Franchin, A., Nieminen, T., Riccobono, F., Hutterli, M., Duplissy, J., Almeida, J., Amorim, A., Breitenlechner, M., Downard, A. J., Dunne, E. M., Flagan, R. C., Kajos, M., Keskinen, H., Kirkby, J., Kupc, A., Kürten, A., Kurtén, T., Laaksonen, A., Mathot, S., Onnela, A., Praplan, A. P., Rondo, L., Santos, F. D., Schallhart, S., Schnitzhofer, R., Sipilä, M., Tomé, A., Tsagkogeorgas, G., Vehkamäki, H., Wimmer, D., Baltensperger, U., Carslaw, K. S., Curtius, J., Hansel, A., Petäjä, T., Kulmala, M., Donahue, N. M., and Worsnop, D. R.: Molecular understanding of atmospheric particle formation from sulfuric acid and large oxidized organic molecules, *Proc. Natl. Acad. Sci.*, 110, 17223–17228, doi:10.1073/pnas.1306973110, 2013.
- Tanner, D. J., Jefferson, A., and Eisele, F. L.: Selected ion chemical ionization mass spectrometric measurement of OH, *J. Geophys. Res.-Atmos.*, 102, 6415–6425, doi:10.1029/96JD03919, 1997.
- Tuazon, E. C., Atkinson, R., Aschmann, S. M., and Arey, J.: Kinetics and products of the gas-phase reactions of O₃ with amines and related compounds, *Res. Chem. Intermed.*, 20, 303–320, 1994.
- Viggiano, A. A. and Arnold, F.: Stratospheric sulfuric acid vapor: New and updated measurements, *J. Geophys. Res.-Ocean.*, 88, 1457–1462, doi:10.1029/JC088iC02p01457, 1983.
- Voigtländer, J., Duplissy, J., Rondo, L., Kürten, A., and Stratmann, F.: Numerical simulations of mixing conditions and aerosol dynamics in the CERN CLOUD chamber, *Atmos. Chem. Phys.*, 12, 2205–2214, doi:10.5194/acp-12-2205-2012, 2012.
- Wimmer, D., Lehtipalo, K., Franchin, A., Kangasluoma, J., Kreissl, F., Kürten, A., Kupc, A., Metzger, A., Mikkilä, J., Petäjä, T., Riccobono, F., Vanhanen, J., Kulmala, M., and Curtius, J.: Performance of diethylene glycol-based particle counters in the sub-3 nm size range, *Atmos. Meas. Tech.*, 6, 1793–1804, doi:10.5194/amt-6-1793-2013, 2013.
- Zhang, R., Suh, I., Zhao, J., Zhang, D., Fortner, E. C., Tie, X., Molina, L. T., and Molina, M. J.: Atmospheric New Particle Formation Enhanced by Organic Acids, *Science* 304, 1487–1490, 2004.