

Technical University of Denmark



Quantification of NS1 dengue biomarker in serum via optomagnetic nanocluster detection

Antunes, Paula Soares Martins; Watterson, Daniel; Parmvi, Mattias; Burger, Robert; Boisen, Anja; Young, Paul; Cooper, Matthew A; Hansen, Mikkel Foug; Ranzoni, Andrea; Donolato, Marco

Published in:
Scientific Reports

Link to article, DOI:
[10.1038/srep16145](https://doi.org/10.1038/srep16145)

Publication date:
2015

Document Version
Publisher's PDF, also known as Version of record

[Link back to DTU Orbit](#)

Citation (APA):
Antunes, P. S. M., Watterson, D., Parmvi, M., Burger, R., Boisen, A., Young, P., ... Donolato, M. (2015). Quantification of NS1 dengue biomarker in serum via optomagnetic nanocluster detection. *Scientific Reports*, 5, 16145. DOI: 10.1038/srep16145

DTU Library

Technical Information Center of Denmark

General rights

Copyright and moral rights for the publications made accessible in the public portal are retained by the authors and/or other copyright owners and it is a condition of accessing publications that users recognise and abide by the legal requirements associated with these rights.

- Users may download and print one copy of any publication from the public portal for the purpose of private study or research.
- You may not further distribute the material or use it for any profit-making activity or commercial gain
- You may freely distribute the URL identifying the publication in the public portal

If you believe that this document breaches copyright please contact us providing details, and we will remove access to the work immediately and investigate your claim.

SCIENTIFIC REPORTS



OPEN

Quantification of NS1 dengue biomarker in serum *via* optomagnetic nanocluster detection

Received: 16 July 2015

Accepted: 07 October 2015

Published: 05 November 2015

Paula Antunes¹, Daniel Watterson³, Mattias Parmvi¹, Robert Burger¹, Anja Boisen¹, Paul Young³, Matthew A. Cooper², Mikkel F. Hansen¹, Andrea Ranzoni² & Marco Donolato¹

Dengue is a tropical vector-borne disease without cure or vaccine that progressively spreads into regions with temperate climates. Diagnostic tools amenable to resource-limited settings would be highly valuable for epidemiologic control and containment during outbreaks. Here, we present a novel low-cost automated biosensing platform for detection of dengue fever biomarker NS1 and demonstrate it on NS1 spiked in human serum. Magnetic nanoparticles (MNPs) are coated with high-affinity monoclonal antibodies against NS1 *via* bio-orthogonal Cu-free 'click' chemistry on an anti-fouling surface molecular architecture. The presence of the target antigen NS1 triggers MNP agglutination and the formation of nanoclusters with rapid kinetics enhanced by external magnetic actuation. The amount and size of the nanoclusters correlate with the target concentration and can be quantified using an optomagnetic readout method. The resulting automated dengue fever assay takes just 8 minutes, requires 6 μ L of serum sample and shows a limit of detection of 25 ng/mL with an upper detection range of 20000 ng/mL. The technology holds a great potential to be applied to NS1 detection in patient samples. As the assay is implemented on a low-cost microfluidic disc the platform is suited for further expansion to multiplexed detection of a wide panel of biomarkers.

Dengue fever is one of the major public health concerns in urban and semi-urban areas across tropical and subtropical regions¹. Over the past decades, it has emerged as the most important mosquito-borne viral disease, spreading globally with a 30-fold increase in incidences. Nowadays, it is endemic in more than 100 countries with half of the world population at risk of infection². The disease is transmitted by infected *Aedes aegypti* mosquitoes and can be divided into four distinct serotypes (DENV 1-4). Undifferentiated symptoms, such as fever, appear rapidly after infection³ and it is therefore critical to provide reliable diagnosis in early stages. Viral culture or nucleic acid amplification offer sufficient accuracy and specificity but they are rarely available for decentralized diagnostics in dengue endemic countries⁴. Serological assays are routinely used to confirm potential infections but are, however, less specific⁵. Additionally, the immune response produces immunoglobulins only in later stages of infection and the anamnestic response in secondary infections complicates interpretation of results⁶. Thus, there is a need for sensitive biomarker-based detection technologies offering early and highly specific detection of dengue fever. One of the most established early biomarkers of dengue fever is non-structural protein NS1, released into the bloodstream during viral replication in moderately high concentrations (up to μ g/mL levels in acute cases)⁷.

¹Department of Micro- and Nanotechnology, Technical University of Denmark, DTU Nanotech, Building 345 East, DK-2800 Kgs. Lyngby, Denmark. ²Institute for Molecular Bioscience, The University of Queensland, 306 Carmody Road, Brisbane, 4072 QLD, Australia. ³School of Chemistry and Molecular Biosciences, The University of Queensland, Brisbane, 4072 QLD, Australia. Correspondence and requests for materials should be addressed to A.R. (email: a.ranzoni@gmail.com) or M.D. (email: marco.donolato@gmail.com)

The proven clinical relevance of early NS1 detection has stimulated the development of immuno-chromatographic lateral flow assays⁸, which are rapid immunoassays (15–20 min) designed to provide a non-quantitative readout at the point-of-care (PoC)⁹. However, in many cases the virus serotype and the infection status of patients limit the sensitivity and reliability of these tests¹⁰ and laboratory confirmation is often required¹¹. Enzyme-linked immunosorbent assays (ELISAs) remain the gold standard in dengue endemic areas but the test can take several hours and requires specialized personnel and laboratory facilities¹².

In response to these challenges several groups have proposed biosensor technologies for NS1 quantification in formats compatible with decentralized diagnostics. Immunosensors based on immunospot assays using fluorescent nanoparticles¹³, surface plasmon resonance¹⁴, and electrochemical detection^{15,16} have recently been presented. These technologies display a growing capacity to provide sensitive NS1 quantification. However, they require multi-step assay strategies and cannot easily be scaled to simultaneous detection of multiple biomarkers. The challenging integration therefore limits their potential for dengue diagnostics³.

Here we present a novel optomagnetic lab-on-a-disk technology for NS1 detection based on aggregation of magnetic nanoparticles (MNPs). Previous validation of the readout principle on a model molecular assay in buffer¹⁷ is now extended to a one-step MNP-based homogeneous immunoassay directly in serum. A biomarker-dependent aggregation of magnetic nanoparticles in raw biological samples is very challenging as nonspecific aggregation cannot be reduced via enhanced stringency of washing steps. Endogenous proteins bind non-specifically and may thus interfere with specific recognition of the target biomarker and impair assay sensitivity. To overcome these challenges, we have designed an anti-fouling surface attachment for the antibodies by means of ‘click’ chemistry¹⁸. The passivated nanoparticles are deployed in a magnetic agglutination assay, where a few microliters of serum sample are mixed with two identical populations of MNPs functionalized, respectively, with capture (Gus11) and reporter (1H7.4) monoclonal antibodies (mAb) raised against NS1 protein. Sample incubation in a strong magnetic field (hereafter named “magnetic incubation”) induces NS1-mediated MNP aggregation. As a final step, the concentration of the target analyte in solution is quantified by measuring the modulation of the transmitted light upon a magnetic field actuation of the nanoclusters¹⁹. The entire assay protocol has been implemented on a disc-based platform, which is suited for inclusion of blood-serum separation and for further future expansion to detect a panel of serological markers. We optimize key assay parameters (MNP concentration, incubation conditions, and sample volume) to achieve a clinically relevant NS1 sensitivity range. Ultimately, we present a dose-response curve directly in serum proving robust NS1 quantification in 8 minutes using a serum volume of only 6 μ L. The lower limit of detection is established to 25 ng/mL and the sensitivity range of NS1 extends up to 20000 ng/mL.

Results and Discussion

MNPs coated with capture (Gus11) and reporter (1H7.4) antibodies bind different epitopes on the NS1 antigen and, for this reason, the presence of NS1 causes linking and agglutination of capture and reporter MNPs with kinetics that can be accelerated by a magnetic incubation step (Fig. 1a). During the magnetic incubation, the strong applied magnetic field causes formation of MNP chains. While in close proximity, the MNPs are subject to Brownian motion, which randomizes their relative orientation and further promotes inter-nanoparticle binding²⁰. The formed clusters are subsequently detected by applying an oscillating uniaxial magnetic field along the path of the laser light: the controlled movements of the formed clusters modulates the transmitted light intensity either via a rotation of individual MNPs or nanoclusters with an optical anisotropy, or via a reversible formation of chains of MNPs along the magnetic field. The MNPs used in this work have a negligible remnant magnetic moment and are substantially spherical. Therefore, they do not physically rotate in an oscillating uniaxial magnetic field and there is no link between the induced magnetic moment in the particles and the optical transmissivity. Moreover, in the weak magnetic field applied during the measurements the magnetic dipole interaction between two particles is much smaller than the thermal energy and therefore chain formation due to magnetic dipole interactions is not energetically favorable. Hence, a homogeneous suspension of non-aggregated particles is not expected to produce a modulation of the transmitted light intensity. This agrees well with our observations on uncoated particle suspensions where the particles are kept well separated by electrostatic repulsion (data not shown). However, when particles form dimers or larger agglomerates, there is a linked magnetic and optical anisotropy that produces a modulation of the transmitted light during a cycle of the magnetic field²¹. When the field magnitude is large, the structures tend to align with their long axis along the magnetic field, whereas their orientation becomes random, due to thermal agitation, when the field is low, i.e., the transmitted light intensity peaks twice during a cycle of the magnetic field. Accordingly, the light modulation occurs at $2f$ and the response can be determined from the 2nd harmonic component of the signal measured from the photodetector. We have previously shown that the complex 2nd harmonic signal can be used to extract information related to the real and imaginary magnetic susceptibility of the MNPs and that the in-phase 2nd harmonic signal, V'_2 , versus frequency of applied field presents a characteristic peak related to the Brownian relaxation frequency of MNP nanoclusters, which is inversely proportional to the hydrodynamic volume of the nanocluster^{17,22}.




Figure 1. (a) Principle of NS1 agglutination assay using functionalized MNPs coated with capture (Gus11) and reporter (1H7.4) monoclonal antibodies. The NS1 target captured by the MNPs (left panel) triggers the formation of MNP nanoclusters during the magnetic incubation (right panel). (b) Schematic of readout system consisting of a blue laser light source (from a Blu-ray optical pickup unit, OPU), a photodetector, and two electromagnets that produce the magnetic field modulating the nanocluster rotation and therefore the light transmission. Opposing the detection circuit are placed two permanent magnets to perform the magnetic incubation. (c) Spectra of the normalized 2nd harmonic photodetector signal vs. frequency of the magnetic field excitation (V_2' vs. f) for an NS1 negative (blank) and an NS1 positive sample.

To illustrate the readout method, Fig. 1c shows the in-phase part of the normalized optomagnetic spectra (V_2' vs. f) for two MNP samples. The blank sample contains MNPs coated with capture and reporter antibodies, respectively, mixed in serum. This sample exhibits a peak at a frequency of $f=17$ Hz of weak intensity, which is attributed to small population of MNP clusters formed via non-specific interactions during the antibody conjugation and during the magnetic incubation step. The peak position corresponds to a hydrodynamic size compatible with the manufacturer specification¹⁹. The NS1 positive sample, where NS1 links MNPs with capture and reporter antibodies (cf. Fig. 1a), shows a spectrum with higher intensity and a slight peak shift to low frequencies due to the formation of specific MNP nanoclusters. The number and size of these MNP nanoclusters increase with the NS1 concentration for the investigated range of concentrations to generate signals of increasing intensity. A dose-response curve is extracted by measuring the value of V_2' at a frequency of $f=17$ Hz, corresponding to the position of the peak at low NS1 concentrations.

To mitigate the formation of nonspecific aggregates during the magnetic incubation and the influence of complex biological fluids, we have designed an anti-fouling magnetic surface architecture based on generation of a monolayer of blocking proteins (human serum albumin, HSA) on the surface of the nanoparticles. The protein monolayer was used to anchor affinity probes by means of bio-orthogonal Cu-free cycloaddition (Fig. 2a)²³. We have previously demonstrated that bio-orthogonal conjugation approaches can dramatically impact assay performances in complex matrices¹⁸. In particular, coating of the hydrophobic regions of polystyrene magnetic nanoparticles with HSA, by means of carbodiimide chemistry helped minimize the adsorption of endogenous interfering agents present in human serum. As it is well known²⁴ that the acylisourea intermediate is highly unstable in aqueous conditions, the molar excess of HSA required to fully saturate the carboxyl groups available on the magnetic nanoparticles had to be carefully controlled. Figure 2b shows the measured number of HSA molecules per MNP as function of the mass of HSA used per mass of the MNPs. It is observed that the number of




Figure 2. (a) Surface molecular architecture: an HSA monolayer (black blocks) on the nanoparticle surface generates a blocking layer onto which bio-orthogonally linked affinity probes are immobilized. (b) Number of HSA molecules per MNP as function of the HSA concentration used to identify the amount of HSA to generate a full protein layer on the MNP surface. (c) Absorption spectra of Gus11 antibody before modification (light orange dashed curve) and after reaction with a DBCO linker in 10-fold molar excess (light orange curve) and 20-fold molar excess (red curve). The degree of modification is estimated from the characteristic peak at 310 nm. (d) Number of immobilized antibodies (Gus11/1H7.4) per MNP estimated using anti-mouse HRP-conjugate antibody assay.

HSA molecules immobilized on the nanoparticle surface progressively increased with the input of HSA. The available surface area was completely occupied by HSA when the reaction occurred in presence of more than $100\mu\text{g}$ HSA per mg of nanoparticles. This closely correlates with the value required for a theoretical monolayer²⁴. The ligation of HSA molecules on the MNP surface introduced primary amines that were subsequently used as handles for the subsequent layer of the molecular surface architecture. Heterofunctional amine-reactive linkers were covalently bound to HSA to introduce azide moieties on the nanoparticle surface. Similarly, azide-complementary dibenzylcyclooctyne (DBCO) moieties were introduced on the antibodies. The DBCO moiety has a unique spectral fingerprint, with an absorption peak at 310 nm that enabled the degree of modification to be directly monitored by measuring UV absorption of the modified biomolecules. Figure 2c shows the absorption spectra obtained for unmodified Gus11 antibodies and antibodies modified with DBCO linker in $10\times$ and $20\times$ molar excess. We carefully controlled the reaction conditions to ensure to introduce 2-3 DBCO molecules per antibody and verified that the modification did not affect the affinity of the antibody to the NS1 protein¹⁸. After 'click' between azide and DBCO, approximately 2000 antibodies were flexibly linked to the HSA monolayer per particle by means of hydrophilic linkers (Fig. 2d).

We first validated the efficacy of the magnetic incubation step by investigating the kinetics of the MNP agglutination in the presence of NS1. A blank sample of serum and a positive sample containing NS1 spiked into serum at a concentration of 1000 ng/mL were used. The MNP concentration was kept constant at 0.1 mg/mL . The reaction was monitored for 60 minutes by measuring optomagnetic spectra every five minutes. The V_2' component at $f=17\text{ Hz}$ corresponding to the position of the peak for low NS1 concentrations was monitored vs. time t for both samples, as shown in Fig. 3a. To eliminate the influence




Figure 3. Kinetics of MNP clustering. (a) Measurements of V_2' (17 Hz) vs. time before and after magnetic incubation for the indicated NS1 concentrations. Values were normalized to the first data point. (b) Comparison of Δ -values obtained using NS1 pre-incubation at room temperature (RT) for different times (0 min, 10 min, 30 min) and at 37°C for a time of 30 min. Δ denotes the difference in V_2' (17 Hz) signals after magnetic incubation between the 1000 ng/mL sample and the blank sample.

of sample-to-sample variation and to ease graphical comparison, the measurements were normalized by the first measured point ($t = 0$).

The first 30 minutes correspond to the capture phase where NS1 molecules are directly captured onto the surface of the MNPs. This binding reaction triggers the formation of specific MNP nanoclusters and an increase of the intensity of the V_2' signal, which gives rise to a slope in the 1000 ng/mL sample signal vs. time. The signal increase was moderate in the capture phase since nanocluster formation is diffusion-limited and hindered by electrostatic repulsion between the MNPs. The formation of non-specific clusters in the blank sample (with zero NS1 concentration) takes place on a much longer time scale and in 30 min the signal increase is negligible. Subsequently, the magnetic incubation step was performed for 60 cycles (180 s) to speed up the agglutination kinetics between the MNPs.

The magnetic incubation consisted of two steps: first sample incubation between the two permanent magnets to enhance formation of clusters and then a shaking step to break up non-specific MNP nanoclusters and enable re-orientation of the MNPs. The overall magnetic incubation procedure strongly accelerated the capture dynamics of NS1 as well as the nanocluster formation (Fig. 3a). This result is well in line with previous literature which shows that the application of continuous²⁵ or pulsed²¹ magnetic field efficiently enhances MNP agglutination. The data in Fig. 3a show that magnetic incubation provides a 3-fold increase of the signal amplitude in the positive sample compared to the blank sample. Non-specific interactions mediated by magnetically induced collisions, physisorbed biomolecules from the human serum and residual hydrophobic regions (e.g., triazole groups) could occur, even in the presence of a HSA monolayer. We observed a net decrease in the signal amplitude during the subsequent 30 minutes for both the 1000 ng/mL and blank samples. Since this effect is target independent, we attribute the decrease to the loss of weakly bound nanoclusters formed due to non-specific interactions.

To maximize the biomarker depletion during the incubation phase, we have evaluated the assay performance as function of temperature and time. Kinetic theory dictates that increased thermal energy results in higher motility and collision rate between biomolecules in solution. In addition, the antibody association-dissociation constant depends on temperature²⁶. We define Δ as the difference between the V_2' (17 Hz) signals for the positive sample and the blank sample just after magnetic incubation. Figure 3b shows Δ measured for a 1000 ng/mL sample with four different incubation conditions. The MNP concentration, the sample volume and the magnetic incubation protocol were kept constant and the same as in the previous experiment. The results indicate that a higher or similar value of Δ is obtained at RT compared to 37°C. Moreover, pre-incubation prior to the magnetic incubation is observed to only result in minor signal improvement compared to no pre-incubation. Hence, we can conclude that the NS1 capture and MNP clustering mainly take place during the magnetic incubation. Other experimental conditions, such as a two-step approach with capture and reporter nanoparticles incubating with the sample in successive steps, were shown to provide a lower Δ (Supplementary Information, Section S2).

We further investigated the parameters of the magnetic incubation. We found that prolonged exposure to the magnetic field resulted focusing of the MNPs onto one side of the microfluidic chamber due to the field not being completely homogeneous. For a magnetic field exposure time of 1 s, we achieved a




Figure 4. $V_2'(17\text{ Hz})$ signal vs. NS1 concentration for the indicated number of magnetic incubation cycles (each lasting 3 s in total). Each protocol was applied for the indicated concentrations of NS1 spiked in serum (total sample used $6\mu\text{L}$) and a MNP concentration of 0.1 mg/mL . Lines are curve fits obtained using the four parameter logistic model as described in the Supplementary Information, Section S3.

dynamic cloud of MNPs perfusing the microfluidic chamber. The shaking step time, acceleration and speed needed to optimally redistribute the particles were optimized empirically. The optimal magnetic incubation protocol length for the NS1 assay was identified by varying the total number of magnetic incubation cycles. Figure 4 shows $V_2'(17\text{ Hz})$ vs. NS1 concentration spiked in serum (total sample volume $6\mu\text{L}$) measured after the indicated number of magnetic incubation cycles. In all cases, MNPs were used at a concentration of 0.1 mg/mL .

Without magnetic incubation, the signal was low and identical for all NS1 concentrations. After magnetic incubation, the signal increased by a factor of 3–10 with a higher increase for higher NS1 concentrations. For all investigated NS1 concentrations, the signal amplitude was higher for the longest magnetic incubation, which causes as well higher signal in the blank sample (0 ng/mL sample). The lines in Fig. 4 are curve fits obtained using the four parameter logistic (4PL) model with the parameters given in Table S1 (Supplementary Information, Section S3). Analysis in terms of the 4PL model supports that 180 cycles of magnetic incubation provides the highest sensitivity. In addition, the highest $V_2'(17\text{ Hz})$ amplitude ratio between the 100 ng/mL sample and the blank after 180 cycles also indicate that this number of cycles results in a lower LOD. To further optimize the assay conditions, the effect of the MNP concentration was studied (Supplementary Information, Section S4) and the optimal MNP concentration was established to 0.1 mg/mL .

The performance of the proposed technology was evaluated using the above optimal assay protocol on NS1 spiked into a serum sample with a total volume of $6\mu\text{L}$. This corresponds to the typical volume that can be extracted out of a single drop of blood from, for example, a patient fingerprick. Figure 5a shows V_2' spectra obtained for the indicated NS1 concentrations. The spectra show a clear signal increase for increasing NS1 concentration with a peak exhibiting a slight shift towards lower frequencies for high NS1 concentrations. For the high investigated NS1 concentrations, multiple NS1 molecules are likely to bind to each MNP (an NS1 concentration of 10000 ng/mL corresponds to 500 NS1 molecules per MNP). Accordingly, larger clusters of MNPs are expected to form with a correspondingly larger hydrodynamic size and lower Brownian relaxation frequency.

Figure 5b shows $V_2'(17\text{ Hz})$ vs. NS1 concentration obtained from triplicate experiments. The solid line shows a curve fit of the 4PL model (Supplementary Information, Section S3) to the data. The Hill coefficient value for this curve was 1.03 and the EC_{50} value was 566 ng/mL . The coefficient of variation did not exceed 2% for any of the measured samples. This indicates that the method is also robust in serum. The limit of detection (LOD), calculated using the blank signal plus three standard deviations, was 25 ng/mL . The dose-response curve showed a large dynamic range covering almost three orders of magnitude of NS1 concentrations. Saturation is reached at high concentrations partly due to the fact that large MNPs are spun on the side of the chamber during magnetic incubation, which is confirmed by the absence of further peak shift. If needed, the linear range could be tailored to, e.g., higher concentrations, by tuning the magnetic incubation protocol and the MNP concentration.

The presented assay dynamic range and LOD are comparable to those obtained using other sensing technologies based on MNP clustering^{21,25,27–29}. Thanks to the efficient antibody coverage on the MNPs surface a comparatively larger dynamic range is obtained in comparison to other approaches where magnetic field actuation is used as well to enhance particle agglutination^{21,29}. A lower limit of detection may potentially be reached in an MNP clustering assay with a miniaturized nuclear magnetic resonance (NMR) based readout^{28,30,31}. However, a simple optical readout, compatible with any optical transparent microfluidic system and which does not require any sophisticated electronic components or complex




Figure 5. (a) V_2' spectra of 0.1 mg/mL MNP suspension mixed with the indicated concentrations of NS1 in serum. Lines are guides to the eye. (b) $V_2'(17 \text{ Hz})$ vs. NS1 concentration in serum. Error bars are standard deviations obtained from triplicate experiments. The line is a curve fit obtained using the four parameter logistic model. The horizontal line indicates the blank sample signal plus 3 standard deviations.

cartridge manufacturing³² holds a stronger potential to impact POC diagnostics of infectious diseases. It is indeed important to put these results into perspective for the detection of *flavivirus* induced diseases, like dengue fever, which predominantly occur in remote areas with minimal access to well-equipped clinical laboratories. This particular diagnostic niche is dominated by ELISA and lateral flow devices due to their cost-effectiveness. ELISA delivers sensitive detection at the expense of a time consuming multi-step procedure whilst the latter offers simplicity with moderate analytical/clinical sensitivity³. Our approach delivers a performance comparable to ELISA, but with a much simpler assay protocol and a much shorter total assay time on a lower sample volume. Moreover, the integration with centrifugal microfluidics enables full automation in an out-of-lab setting and is anticipated to lead to future multiplexed detection from the same drop of blood.

The levels of circulating NS1 in dengue infected individuals can vary depending on the severity of the disease, its immunological status (primary or secondary infection) and its serotype. In acute cases, circulating NS1 values in the $\mu\text{g/mL}$ range have been reported⁷. However, values below 10 ng/mL may occur for early stages of the disease³³. Current ELISA commercial kits, the gold standard for NS1 detection, only partially cover this low range of concentrations and poor clinical sensitivities for DENV-2 and DENV-4 serotypes have been reported³⁴. An improved LoD would be beneficial for early diagnosis of dengue. The presented technology has the potential to deliver rapid and fully automated diagnosis in the field. The main factor limiting the current LOD is the nonspecific binding of MNPs. Further optimization of the MNP surface architecture and the assay protocol are likely to reduce the non-specific background. Moreover, more sophisticated analysis methods, for example, taking into account the full measured frequency spectra before and after magnetic incubation rather than the value at a single frequency, may also improve the sensitivity and LOD. These improvements are essential for translation to the clinics and usage in the field and are topics of ongoing work. We are presently developing novel affinity probes capable of cross-reactive interaction across all four dengue serotypes. These will enable estimation of the clinical sensitivity and specificity by testing samples of dengue infecting patients. Precise quantification of other biomarkers, including serological markers, could help triaging patients and may further provide valuable data for surveillance and epidemic control.

In conclusion, we have presented a novel biosensing platform for dengue NS1 detection spiked in human serum. The assay is fully integrated and requires only a few microliters of clinical sample to be mixed with the bioactive nanoparticles. External magnetic fields actuate the nanoparticles in solution, triggering both biomarker-induced nanocluster formation and a time-dependent transmitted light modulation profile, which correlates with the number of nanoclusters in solution. We have demonstrated minimal non-specific nanoparticle aggregation by coating nanoparticles with an anti-fouling surface molecular architecture. Bio-orthogonal 'Cu-free' click chemistry enables efficient immobilization of >2000 high affinity antibodies against dengue NS1 protein. The result is a fast (8 minutes) and automated dengue fever diagnostic assay compatible with a fingerprick sample volume. The presented technology shows a limit of detection of 25 ng/mL and a wide dynamic range up to 20000 ng/mL.

Materials and Methods

Chemicals and materials. The generation and characterization of 1H7.4 and Gus11 antibodies have been described previously^{11,14,35,36}. In brief the antibodies were produced by immunizing BALB/c mice with subcutaneous injections of NS1 protein, regularly boosted with subsequent monthly injections.

Blood samples were collected by tail bleed and the antibody titer was checked by ELISA to identify the mice with the strongest immune response prior to culling. The harvested spleen cells were fused with plasmacytoma cell line SP2/0 at a ratio 10:1 and subsequent selection and characterization of the antibodies was performed by ELISA. To generate purified stocks, antibodies were isolated from ascites fluid using affinity chromatography on protein G sepharose (HiTrap Protein G HP, GE Healthcare) as per manufacturer instructions. After elution in 100 mM glycine pH 2.7, purified antibody stocks were buffer exchanged using centrifugal filtration (50kDa MWCO Amicon Ultra-15, Millipore) into phosphate buffer saline (PBS). Protein concentrations were determined using the Bicinchoninic acid (BCA) Protein Assay (Pierce).

Superparamagnetic polystyrene nanoparticles with embedded magnetic nanograins and a nominal diameter of 170 nm and COOH surface were acquired from Merck (NJ, USA). Purified recombinant Dengue Virus NS1 glycoprotein was obtained from Hawaii Biotech (Lot. 2DNS1). All chemicals were purchased from Sigma Aldrich, unless otherwise noted. MNP storage buffer was prepared with PBS and Tween 20. In this investigation, NS1 was always diluted in CRP-free serum from HyTest Ltd. (Turku, Finland).

The methods were carried out in accordance with the approved guidelines. All experimental protocols were approved by the DTU Nanotech work environment committee. The procedure for coupling of Gus11 and 1H7.4 antibodies to separate populations of MNPs is described in the Supplementary Information, Section S1. Unless otherwise stated, all MNP suspensions were 1:1 mixtures of MNPs with Gus11 and 1H7.4 antibodies and stated MNP concentrations refer to the mixture of MNPs.

Experimental Setup. Figure 1b shows a picture of the readout system previously described in refs 17 and 22. In brief, the major optical elements were a Sanyo Blu-ray optical pick-up unit ($\lambda = 405$ nm) and a ThorLabs PDA36A photodetector. The microfluidic disc was placed between two electromagnets used to generate a sinusoidal uniaxial magnetic field parallel to the laser beam direction of fixed amplitude $B_0 = 2$ mT at a frequency f up to 300 Hz. The pre-amplified signal from the photodetector was acquired using a National Instruments 6251 data acquisition (DAQ) card and analyzed through a software-based lock-in amplifier. The same DAQ card controlled a customized circuit, embedded in the optical pickup unit control board, used to provide the alternating current to the electromagnets. The software extracted the intensity of the 2nd harmonic signal from the photodetector and its phase lag with respect to the magnetic field excitation and presented it as the complex 2nd harmonic lock-in signal. The software also extracted the average signal from the photodetector. All signals presented below have been normalized with the average photodetector signal and are hence dimensionless. The normalized in-phase 2nd harmonic photodetector signal is given the symbol V_2' .

The disc was connected to a rotary stepper motor (Maxxon Motor, mod. 273756, Switzerland) used to perform the microfluidic operations. The motor was operated using a digital positioning controller (Maxxon, mod. 347717) via a custom-made Labview program (National Instruments, US). A pair of permanent magnets was placed on the radial position opposite to the two electromagnets and was used to provide a strong approximately homogeneous field during the magnetic incubation step. The field measured at the center of the measurement chamber was around 90 mT.

Disc fabrication. Centrifugal microfluidic discs were manufactured using Poly(methyl methacrylate) (PMMA, Axxicon, Netherlands) and pressure sensitive adhesive (PSA, 90106, Adhesives Research, Ireland) as described by Donolato *et al.*, 2014. Each microfluidic disc was made from three 600 μ m thick PMMA DVD halves. Reservoirs were created by cutting through the first PMMA disc using a CO₂ laser (Mini 18, 30W, Epilog, USA). Inlets were laser machined in the second disc and the third disc was used as support. Alignment holes were also cut in all three discs. Following these steps, fluid channels were created in two PSA sheets using a blade cutter machine (Silhouette, USA). The PSA was laminated on the bottom and top discs, which were then aligned and bonded to the central disc containing the reservoirs. This rapid prototyping process allowed for the complete fabrication of a disc with eighteen microfluidic chambers in less than 20 min.

Measurement protocol. Experiments were performed as follows: each MNP population was diluted in the proper buffer to the required concentration followed by vortex and ultrasound treatment for 2 min (Model 150 V, Biologics, US). NS1 was spiked in human serum and aliquoted accordingly. Samples were prepared off-disc by mixing the MNP suspension with 6 μ L of NS1 sample and immediately transferred into the disc loading chamber. The disc was then placed on the previously described setup and spun (18 Hz, 3 sec) to drive the sample to the detection chamber. The chamber was aligned with the Blu-ray laser and the two electromagnets, and a first optomagnetic spectrum was recorded. An optomagnetic spectrum consisted of a measurement of V_2' vs. f for fifteen logarithmically equidistant values of f between 1 and 300 Hz and it was recorded in 60 s. Subsequently, the sample was exposed to a magnetic incubation step. The magnetic incubation protocol comprised a number of cycles, each with 1 sec of incubation between the permanent magnets followed by 2 sec of mixing (one full accelerating revolution in each rotation direction). The first step enhanced magnetic agglutination of the MNPs. The second step promoted mixing of the MNPs, random MNP rotation, and random antibody-antigen encounters to

accelerate the recognition²¹. Magnetic incubation protocols with 60, 120 and 180 cycles were investigated. Finally, after magnetic incubation, the chamber was aligned with the laser to record the second optomagnetic spectrum.

References

1. WHO. Global Strategy for Dengue Prevention and Control 2012–2020. 35 (2012).
2. Bhatt, S. *et al.* The global distribution and burden of dengue. *Nature* **496**, 504–507 (2013).
3. Tang, K. F. & Ooi, E. E. Diagnosis of dengue : an update. *Expert Rev Anti Infect Ther* **10**, 895–907 (2012).
4. Wilder-Smith, A. *et al.* Dengue tools: Innovative tools and strategies for the surveillance and control of dengue. *Glob. Health Action* **5**, 1–9 (2012).
5. Peeling, R. W. *et al.* Evaluation of diagnostic tests: dengue. *Nat. Rev. Microbiol.* **8**, S30–S37 (2010).
6. Kao, C.-L., King, C.-C., Chao, D.-Y., Wu, H.-L. & Chang, G.-J. J. Laboratory diagnosis of dengue virus infection: current and future perspectives in clinical diagnosis and public health. *J. Microbiol. Immunol. Infect.* **38**, 5–16 (2005).
7. Alcon, S. *et al.* Enzyme-linked immunosorbent assay specific to dengue virus type 1 nonstructural protein NS1 reveals circulation of the antigen in the blood during the acute phase of disease in patients experiencing primary or secondary infections. *J. Clin. Microbiol.* **40**, 376–381 (2002).
8. Wang, S. M. & Sekaran, S. D. Early diagnosis of dengue infection using a commercial dengue duo rapid test kit for the detection of NS1, IGM, and IGG. *Am. J. Trop. Med. Hyg.* **83**, 690–695 (2010).
9. Tricou, V. *et al.* Comparison of two dengue NS1 rapid tests for sensitivity, specificity and relationship to viraemia and antibody responses. *BMC Infect. Dis.* **10**, 142 (2010).
10. Duong, V. *et al.* Clinical and virological factors influencing the performance of a NS1 antigen-capture assay and potential use as a marker of dengue disease severity. *PLoS Negl. Trop. Dis.* **5**, 1244 (2011).
11. Hang, V. T. *et al.* Diagnostic accuracy of NS1 ELISA and lateral flow rapid tests for dengue sensitivity, specificity and relationship to viraemia and antibody responses. *PLoS Negl. Trop. Dis.* **3**, 1–7 (2009).
12. Huang, C.-H. *et al.* Laboratory diagnostics of dengue fever: an emphasis on the role of commercial dengue virus nonstructural protein 1 antigen rapid test. *J. Microbiol. Immunol. Infect.* **46**, 358–365 (2013).
13. Linares, E. M., Pannuti, C. S., Kubota, L. T. & Thalhammer, S. Immunospot assay based on fluorescent nanoparticles for Dengue fever detection. *Biosens. Bioelectron.* **41**, 180–185 (2013).
14. Hu, D. *et al.* Comparison of surface plasmon resonance, resonant waveguide grating biosensing and enzyme linked immunosorbent assay (ELISA) in the evaluation of a dengue virus immunoassay. *Biosensors* **3**, 297–311 (2013).
15. Silva, M. M. S. *et al.* Electrochemical detection of dengue virus NS1 protein with a poly(allylamine)/carbon nanotube layered immunoelectrode. *J. Chem. Technol. Biotechnol.* **90**, 194–200 (2014).
16. Figueiredo, A. *et al.* Electrical detection of dengue biomarker using egg yolk immunoglobulin as the biological recognition element. *Sci. Rep.* **5**, 7865 (2015).
17. Donolato, M. *et al.* Novel readout method for molecular diagnostic assays based on optical measurements of magnetic nanobead dynamics. *Anal. Chem.* **87**, 1622–1629 (2015).
18. Sanjaya, K. C., Ranzoni, A., Watterson, D., Young, P. & Cooper, M. A. Evaluation of direct versus multi-layer passivation and capture chemistries for nanoparticle-based biosensor applications. *Biosens. Bioelectron.* **67**, 769–774 (2015).
19. Donolato, M. *et al.* Quantification of rolling circle amplified DNA using magnetic nanobeads and a Blu-ray optical pick-up unit. *Biosens. Bioelectron.* **67**, 649–655 (2014).
20. Ranzoni, A., den Hamer, A., Karoli, T., Buechler, J. & Cooper, M. A. Improved immunoassay sensitivity in serum as a result of polymer-entrapped quantum dots: “papaya particles”. *Anal. Chem.* **87**, 6150–6157 (2015).
21. Ranzoni, A., Sabatte, G., van Ijzendoorn, L. J. & Prins, M. W. J. One-step homogeneous magnetic nanoparticle immunoassay for biomarker detection directly in blood plasma. *ACS Nano* **6**, 3134–3141 (2012).
22. Bejhed, R. S. *et al.* Turn-on optomagnetic bacterial DNA sequence detection using volume-amplified magnetic nanobeads. *Biosens. Bioelectron.* **66**, 405–411 (2014).
23. Thirumurugan, P., Matosiuk, D. & Jozwiak, K. Click chemistry for drug development and diverse chemical – biology applications. *Chem. Rev.* **133**, 4905–4979 (2013).
24. Hermanson, G. T. *Bioconjugate Techniques*. 989–1016 (Elsevier, 2013).
25. Baudry, J. *et al.* Acceleration of the recognition rate between grafted ligands and receptors with magnetic forces. *Proc. Natl. Acad. Sci. USA* **103**, 16076–8 (2006).
26. Reverberi, R. & Reverberi, L. Factors affecting the antigen-antibody reaction. *Blood Transfus.* **5**, 227–240 (2007).
27. Issadore, D. *et al.* Magnetic sensing technology for molecular analyses. *Lab Chip* **14**, 2385–97 (2014).
28. Issadore, D. *et al.* Miniature magnetic resonance system for point-of-care diagnostics. *Lab Chip* **11**, 2282–2287 (2011).
29. Daynès, A. *et al.* Fast magnetic field-enhanced linear colloidal agglutination immunoassay. *Anal. Chem.* 150714134125008 (2015). doi: 10.1021/acs.analchem.5b00279
30. Perez, J. M., Josephson, L., O’Loughlin, T., Högemann, D. & Weissleder, R. Magnetic relaxation switches capable of sensing molecular interactions. *Nat. Biotechnol.* **20**, 816–20 (2002).
31. Chung, H. J., Castro, C. M., Im, H., Lee, H. & Weissleder, R. A magneto-DNA nanoparticle system for rapid detection and phenotyping of bacteria. *Nat. Nanotechnol.* **8**, 369–75 (2013).
32. Liang, M. *et al.* Magnetic barcode assay for genetic detection of pathogens. *Nat. Commun.* **4**, 1752 (2013).
33. Allonso, D., Meneses, M. D. F., Fernandes, C. a., Ferreira, D. F. & Mohana-Borges, R. Assessing positivity and circulating levels of NS1 in samples from a 2012 dengue outbreak in Rio de Janeiro, Brazil. *PLoS One* **9**, e113634 (2014).
34. Guzman, M. G. *et al.* Dengue: a continuing global threat. *Nat. Rev. Microbiol.* **8**, S7–16 (2010).
35. Falconar, A. K. I. & Young, P. R. Production of dimer-specific and dengue virus group cross-reactive mouse monoclonal antibodies to the dengue 2 virus non-structural glycoprotein NS1. *J. Gen. Virol.* **72**, 961–965 (1991).
36. Falconar, A. K. I., Young, P. R. & Miles, M. A. Precise location of sequential dengue virus subcomplex and complex B cell epitopes on the nonstructural-1 glycoprotein. *Arch. Virol.* **137**, 315–326 (1994).

Acknowledgements

This work is financially supported by the ERC Advanced grant No. 320535-HERMES, the Danish Strategic Research Council project MUSE and EU FP7 grant No. 604448-NanoMag. M.D. gratefully acknowledges financial support from the Ørsted postdoctoral grant. The authors would also like to acknowledge The University of Queensland for partially funding this work.

Author Contributions

M.D., A.R., M.F.H. and P.A. conceived the study. P.A. and A.R. performed the experiments. P.Y. and D.W. provided the antibodies. M.P. and R.B. provided technical support in the particle functionalization and disk fabrication. A.B. and M.A.C. discussed the results. M.D., A.R., P.A. and M.F.H. wrote the manuscript. All authors reviewed and commented the manuscript.

Additional Information

Supplementary information accompanies this paper at <http://www.nature.com/srep>

Competing financial interests: The authors declare no competing financial interests.

How to cite this article: Antunes, P. *et al.* Quantification of NS1 dengue biomarker in serum *via* optomagnetic nanocluster detection. *Sci. Rep.* **5**, 16145; doi: 10.1038/srep16145 (2015).



This work is licensed under a Creative Commons Attribution 4.0 International License. The images or other third party material in this article are included in the article's Creative Commons license, unless indicated otherwise in the credit line; if the material is not included under the Creative Commons license, users will need to obtain permission from the license holder to reproduce the material. To view a copy of this license, visit <http://creativecommons.org/licenses/by/4.0/>