

12. Status of the IGISOL device

A. Wojtasiewicz¹, G. Cachel², A. Emsallem⁵, M. Gierlik¹, Z. Janas¹, M. Kisieliński,
W. Kurcewicz¹, J. Kurpeta¹, H. Lefort³, S. Lewandowski¹, H. Penttilä⁴, A. Plochocki¹, E. Roeckl¹,
B. Roussi re³, S. Sidor¹ and A. Syntfeld¹

1) *Institute of Experimental Physics, Warsaw University, Poland*

2) *Centre d'Etudes Nucl aires de Bordeaux-Gradignan, France*

3) *Institut de Physique Nucl aire, Orsay, France*

4) *Department of Physics, University of Jyv skyl , Finland*

5) *Institut de Physique Nucl aire, Lyon, France*

After first test experiments [1] a series of improvements was introduced last year at the IGISOL device. The most important of them was a new helium supply and control system and additional vacuum pumps .

The new helium supply and control system consists of high quality stainless steel Swagelok elements and control units for helium flow. Application of this design had considerably decreased the impurities in the helium passing through the helium cell.

Two additional pumps, a primary pump and a turbo one installed at the extraction region of the IGISOL had improved the vacuum about one order of magnitude in the region between the helium cell and the extraction electrode of the mass separator.

These improvements had caused increasing of the efficiency of the gas cell and extraction system about 8 times in comparison with the previous conditions.

In these conditions we continued investigation of a gas catcher/ion guide system. The gas cell (helium chamber) of 120 cm³ volume [2] was filled with high-purity (99,9999%) helium. The α spectra of the recoil nucleus ²¹⁹Rn ($T_{1/2}=3,96$ s) and its daughter ²¹⁵Po ($T_{1/2}=1,78$ ms) were measured by two silicon detectors placed in front of and behind the IGISOL magnet. Extraction efficiencies for ²¹⁹Rn and ²¹⁵Po ions were measured as function of helium pressure in the cell and distance of the ²²³Ra source from the cell exit hole. In order to investigate the so called "plasma effect" the 3,5 MeV/u ¹⁴N beam was passed through the cell for some measurements.

The efficiency of the gas catcher/ion system measured for ²¹⁹Rn was between 25% and 55% depending on the position of the α source and without mass separation. After mass separation the maximum yield is 10%, decreasing with ¹⁴N beam by a factor of ~9 for both ²¹⁹Rn¹⁺ and ²¹⁵Po¹⁺ ions. The estimated extraction time of 2,4 ms is obtained for ²¹⁵Po¹⁺.

In the later measurements the α source (²²³Ra) was removed and the heavy-ion reaction ¹⁴N (5,8 MeV/u) + ²⁰⁹Bi was studied. The short-lived products ²¹³Rn ($T_{1/2}=25$ ms), ²¹⁴Ra ($T_{1/2}=2,5$ s) and ²¹⁵Ra ($T_{1/2}=1,6$ ms) were identified. The observation of the very short-lived isotope ²¹⁵Ra seems to confirm the promising results concerning the high extraction efficiency, the short extraction time and limited influence of the plasma effect [3].

These works were partially performed in the frame of the Warsaw University – IN2P3 (Nr 97-87 and 99-96) and the ION CATCHER (Nr HPRI-CT-2001-50022) collaborations.

References

[1] A.Wojtasiewicz et al., Ann. Rep. HIL Warsaw Univ.(2002) p. 35

[2] A.Wojtasiewicz et al., Ann. Rep. HIL Warsaw Univ.(1998) p. 15

[3] A.Wojtasiewicz et al., RNB6 Conf. ANL Sept. 22-26 (2003) to be publ. in Nucl.Phys. A.